

MI3

Miniature Infrared Sensor



Operating Instructions



Rev. F1 05/2013
55201

Contacts

Raytek Corporation
Worldwide Headquarters
Santa Cruz, CA USA
Tel: +1 800 227 – 8074
(USA and Canada only)
+1 831 458 – 3900
solutions@raytek.com

European Headquarters
Berlin, Germany
Tel: +49 30 4 78 00 80
raytek@raytek.de

France

info@raytek.fr

United Kingdom

ukinfo@raytek.com

Fluke Service Center
Beijing, China
Tel: +86 10 6438 4691
info@raytek.com.cn

Internet: <http://www.raytek.com/>

Thank you for purchasing this Raytek product. Register today at www.raytek.com/register to receive the latest updates, enhancements and software upgrades!



© Raytek Corporation.

Raytek and the Raytek Logo are registered trademarks of Raytek Corporation.

All rights reserved. Specifications subject to change without notice.

WARRANTY

The manufacturer warrants this product to be free from defects in material and workmanship under normal use and service for the period of two years from date of purchase, except as hereinafter provided. This warranty extends only to the original purchaser (a purchase from the manufacturer or a licensed distributor of the manufacturer is an original purchase). This warranty shall not apply to fuses, batteries, or any product which has been subject to misuse, neglect, accident, or abnormal conditions of operation.

In the event of failure of a product covered by this warranty, the manufacturer will repair the instrument when it is returned by the purchaser, freight prepaid, to an authorized Service Facility within the applicable warranty period, provided manufacturer's examination discloses to its satisfaction that the product was defective. The manufacturer may, at its option, replace the product in lieu of repair. With regard to any covered product returned within the applicable warranty period, repairs or replacement will be made without charge and with return freight paid by the manufacturer, unless the failure was caused by misuse, neglect, accident, or abnormal conditions of operation or storage, in which case repairs will be billed at a reasonable cost. In such a case, an estimate will be submitted before work is started, if requested.

THE FOREGOING WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OF MERCHANTABILITY, FITNESS, OR ADEQUACY FOR ANY PARTICULAR PURPOSE OR USE. THE MANUFACTURER SHALL NOT BE LIABLE FOR ANY SPECIAL, INCIDENTAL OR CONSEQUENTIAL DAMAGES, WHETHER IN CONTRACT, TORT, OR OTHERWISE.

SOFTWARE WARRANTY

The manufacturer does not warrant that the software described herein will function properly in every hardware and software environment. This software may not work in combination with modified or emulated versions of Windows operating environments, memory-resident software, or on computers with inadequate memory. The manufacturer warrants that the program disk is free from defects in material and workmanship, assuming normal use, for a period of one year. Except for this warranty, the manufacturer makes no warranty or representation, either expressed or implied, with respect to this software or documentation, including its quality, performance, merchantability, or fitness for a particular purpose. As a result, this software and documentation are licensed "as is," and the licensee (i.e., the User) assumes the entire risk as to its quality and performance. The liability of the manufacturer under this warranty shall be limited to the amount paid by the User. In no event shall the manufacturer be liable for any costs including but not limited to those incurred as a result of lost profits or revenue, loss of use of the computer software, loss of data, the cost of substitute software, claims by third parties, or for other similar costs. The manufacturer's software and documentation are copyrighted with all rights reserved. It is illegal to make copies for another person.

Specifications subject to change without notice.



The device complies with the requirements of the European Directives.

EC – Directive 2004/108/EC (EMC)

Content

| | |
|--|-----------|
| 1 SAFETY INSTRUCTIONS | 10 |
| 2 DESCRIPTION | 11 |
| 2.1 OVERVIEW COMM BOXES | 12 |
| 3 TECHNICAL DATA | 13 |
| 3.1 MEASUREMENT SPECIFICATION | 13 |
| 3.1.1 Sensing Heads | 13 |
| 3.1.2 Comm Box | 14 |
| 3.1.2.1 Comm Box (metal) | 14 |
| 3.1.2.2 Comm Box (DIN)..... | 14 |
| 3.2 OPTICAL CHARTS | 15 |
| 3.3 ELECTRICAL SPECIFICATION..... | 16 |
| 3.3.1 Comm Box, all models | 16 |
| 3.3.2 Comm Box (metal) | 16 |
| 3.3.3 Comm Box (DIN 6TE, analog) | 16 |
| 3.4 ENVIRONMENTAL SPECIFICATION | 17 |
| 3.4.1 Sensing Head..... | 17 |
| 3.4.2 Comm Box (metal) | 17 |
| 3.4.3 Comm Box (DIN)..... | 18 |
| 3.4.4 LTH Electronics | 18 |
| 3.5 DIMENSIONS | 18 |
| 3.5.1 Sensing Head LT, G5 | 18 |
| 3.5.2 Sensing Head LTH..... | 18 |
| 3.5.3 Sensing Head 1M, 2M..... | 19 |
| 3.5.4 Comm Box (metal) | 19 |
| 3.5.5 Comm Box (DIN)..... | 20 |
| 3.6 SCOPE OF DELIVERY..... | 20 |
| 3.6.1 Sensing Head..... | 20 |
| 3.6.2 Comm Box..... | 20 |
| 4 BASICS | 21 |
| 4.1 MEASUREMENT OF INFRARED TEMPERATURE..... | 21 |
| 4.2 EMISSIVITY OF TARGET OBJECT | 21 |
| 4.3 AMBIENT TEMPERATURE..... | 21 |
| 4.4 ATMOSPHERIC QUALITY..... | 21 |
| 4.5 ELECTRICAL INTERFERENCE..... | 21 |
| 5 INSTALLATION | 23 |
| 5.1 POSITIONING..... | 23 |
| 5.1.1 Distance to Object | 23 |
| 5.2 INSTALLATION SCHEMES..... | 23 |
| 5.2.1 Comm Box (metal) | 23 |
| 5.2.2 Comm Box (DIN)..... | 24 |
| 5.3 WIRING, HEAD CABLE | 24 |
| 5.3.1 Comm Box (metal) | 25 |
| 5.3.2 Comm Box (DIN)..... | 25 |
| 5.4 WIRING, TERMINAL | 25 |

| | |
|---|-----------|
| 5.4.1 Comm Box (metal)..... | 26 |
| 5.4.2 Comm Box (DIN 3TE) | 27 |
| 5.4.3 Comm Box (DIN 4 TE) | 28 |
| 5.4.4 Comm Box (DIN 6 TE) | 29 |
| 5.4.5 EMI Resistance for Comm Box (DIN)..... | 32 |
| 5.5 POWER ON PROCEDURE | 33 |
| 5.5.1 One Head System | 33 |
| 5.5.2 Multiple Heads – Random Address Assignment | 33 |
| 5.5.3 Multiple Heads – User Controlled Address Assignment | 33 |
| 5.6 USB..... | 34 |
| 5.7 FIELDBUS | 36 |
| 5.7.1 Addressing | 36 |
| 5.7.2 RS485 based Installations..... | 36 |
| 6 OUTPUTS | 37 |
| 6.1 ANALOG OUTPUT OUT1 | 37 |
| 6.2 ANALOG OUTPUT OUT2 | 37 |
| 6.3 ANALOG OUTPUTS OUT1 - OUT4..... | 38 |
| 6.4 ALARM OUTPUT RELAY..... | 38 |
| 6.5 THERMOCOUPLE OUTPUT TC..... | 39 |
| 7 INPUTS | 40 |
| 7.1 EMISSIVITY (ANALOG)..... | 40 |
| 7.2 EMISSIVITY (DIGITAL)..... | 41 |
| 7.3 AMBIENT TEMPERATURE COMPENSATION..... | 41 |
| 7.4 TRIGGER/HOLD | 43 |
| 7.5 LASER SWITCHING | 44 |
| 8 OPERATION | 45 |
| 8.1 CONTROL PANEL | 45 |
| 8.2 <HEAD> PAGE | 47 |
| 8.3 <BOX SETUP> PAGE | 48 |
| 8.4 <BOX INFO> PAGE | 50 |
| 8.5 POST PROCESSING | 50 |
| 8.5.1 Averaging | 50 |
| 8.5.2 Peak Hold..... | 51 |
| 8.5.3 Valley Hold | 51 |
| 8.5.4 Advanced Peak Hold..... | 52 |
| 8.5.5 Advanced Valley Hold | 53 |
| 8.5.6 Advanced Peak Hold with Averaging..... | 53 |
| 8.5.7 Advanced Valley Hold with Averaging..... | 53 |
| 9 OPTIONS..... | 54 |
| 10 ACCESSORIES | 55 |
| 10.1 ACCESSORIES (ALL MODELS)..... | 55 |
| 10.1.1 Multi-Channel Box..... | 55 |
| 10.1.2 USB/RS485 Adapter | 57 |
| 10.2 ACCESSORIES (LT, G5 HEADS) | 58 |
| 10.2.1 Adjustable Mounting Bracket | 60 |
| 10.2.2 Fixed Mounting Bracket..... | 61 |

| | |
|---|-----------|
| 10.2.3 Air Purge Jacket | 61 |
| 10.2.4 Air Cooling System | 62 |
| 10.2.5 Right Angle Mirror | 66 |
| 10.2.6 Protective Windows | 67 |
| 10.2.7 Close Focus Lens | 67 |
| 10.3 ACCESSORIES (1M, 2M HEADS) | 68 |
| 10.3.1 Fixed Mounting Bracket | 70 |
| 10.3.2 Adjustable Mounting Bracket | 71 |
| 10.3.3 Air Purge Collar | 71 |
| 10.3.4 Right Angle Mirror | 72 |
| 10.3.5 Protective Window | 72 |
| 11 MAINTENANCE | 74 |
| 11.1 TROUBLESHOOTING MINOR PROBLEMS | 74 |
| 11.2 FAIL-SAFE OPERATION | 74 |
| 11.3 CLEANING THE LENS | 75 |
| 11.4 SENSING HEAD EXCHANGE | 76 |
| 12 DATATEMP MULTIDROP SOFTWARE | 77 |
| 12.1 SOFTWARE FEATURES | 77 |
| 12.2 PC REQUIREMENTS | 77 |
| 12.3 USB DRIVER INSTALLATION | 77 |
| 12.4 SOFTWARE LAUNCH | 77 |
| 13 RS485 | 78 |
| 13.1 WIRING | 78 |
| 13.1.1 Comm Box (metal) | 78 |
| 13.1.2 Comm Box (DIN) | 79 |
| 13.2 ASCII PROGRAMMING | 79 |
| 14 PROFIBUS | 80 |
| 14.1 WIRING | 80 |
| 14.1.1 Comm Box (metal) | 80 |
| 14.1.2 Comm Box (DIN) | 82 |
| 14.2 PROGRAMMING | 83 |
| 14.2.1 Parameter Data | 83 |
| 14.2.2 Input Data | 84 |
| 14.2.3 Output Data | 85 |
| 14.2.4 Diagnose Data | 85 |
| 15 MODBUS | 88 |
| 15.1 WIRING | 88 |
| 15.1.1 Comm Box (metal) | 88 |
| 15.1.2 Comm Box (DIN) | 89 |
| 15.2 PROGRAMMING | 90 |
| 15.2.1 Supported Functions | 90 |
| 15.2.2 Parameter Data | 90 |
| 15.2.2.1 Box Parameter | 90 |
| 15.2.2.2 Head Parameter | 92 |
| 16 ETHERNET | 94 |
| 16.1 WIRING | 94 |

| | |
|--|------------|
| 16.1.1 Comm Box (metal) | 94 |
| 16.1.2 Comm Box (DIN) | 94 |
| 16.2 ASCII PROGRAMMING | 94 |
| 16.3 HTTP SERVER | 95 |
| 16.3.1 Data Logging | 96 |
| 17 PROFINET | 97 |
| 17.1 WIRING | 97 |
| 17.1.1 Status LED | 97 |
| 17.2 PROGRAMMING | 98 |
| 17.2.1 I/O Device Configuration | 98 |
| 17.2.1.1 GSD File | 98 |
| 17.2.1.2 Configuration | 98 |
| 17.2.2 Parameter Setting | 98 |
| 17.2.2.1 Station Parameters | 99 |
| 17.2.2.2 Pyrometer Module Parameters | 100 |
| 17.2.3 Input Data Structure | 101 |
| 17.2.3.1 Station Input data | 101 |
| 17.2.3.2 Pyrometer Module Input Data | 101 |
| 17.2.4 Output Data Structure | 101 |
| 17.2.5 Diagnostics | 101 |
| 17.2.5.1 Error Bits of the Station Diagnosis (optional) | 102 |
| 17.2.5.2 Error Bits of the Pyrometer Module Diagnosis (optional) | 102 |
| 18 ASCII PROGRAMMING | 103 |
| 18.1 TRANSFER MODES | 103 |
| 18.2 COMMAND STRUCTURE | 103 |
| 18.3 DEVICE INFORMATION | 104 |
| 18.4 DEVICE SETUP | 104 |
| 18.4.1 Temperature Calculation | 104 |
| 18.4.2 Temperature Pre-Processing | 105 |
| 18.4.3 Emissivity Setting and Alarm Set points | 105 |
| 18.4.4 Post Processing | 106 |
| 18.5 DYNAMIC DATA | 106 |
| 18.6 DEVICE CONTROL | 106 |
| 18.6.1 Output for the Target Temperature | 106 |
| 18.6.2 Analog Output, Scaling | 106 |
| 18.6.3 Alarm Output | 107 |
| 18.6.4 Factory default values | 107 |
| 18.6.5 Lock Mode | 107 |
| 18.6.6 Mode Setting for the Digital Input FTC3 | 107 |
| 18.6.7 Ambient Background Temperature Compensation | 107 |
| 18.7 ADDRESSING OF MULTIPLE HEADS | 107 |
| 18.8 ADDRESSING OF MULTIPLE BOXES (RS485) | 108 |
| 18.9 COMMAND SET | 108 |
| 18.9.1 ASCII Commands for Ethernet and Profinet | 114 |
| 19 APPENDIX | 115 |
| 19.1 DETERMINATION OF EMISSIVITY | 115 |
| 19.2 TYPICAL EMISSIVITY VALUES | 115 |

20 NOTES 119

Safety Instructions

1 Safety Instructions

This document contains important information, which should be kept at all times with the instrument during its operational life. Other users of this instrument should be given these instructions with the instrument. Eventual updates to this information must be added to the original document. The instrument should only be operated by trained personnel in accordance with these instructions and local safety regulations.

Acceptable Operation

This instrument is intended only for the measurement of temperature. The instrument is appropriate for continuous use. The instrument operates reliably in demanding conditions, such as in high environmental temperatures, as long as the documented technical specifications for all instrument components are adhered to. Compliance with the operating instructions is necessary to ensure the expected results.

Unacceptable Operation

The instrument should not be used for medical diagnosis.

Replacement Parts and Accessories

Use only original parts and accessories approved by the manufacturer. The use of other products can compromise the operational safety and functionality of the instrument.

Instrument Disposal



Disposal of old instruments should be handled according to professional and environmental regulations as electronic waste.

Operating Instructions

The following symbols are used to highlight essential safety information in the operation instructions:



Helpful information regarding the optimal use of the instrument.



Warnings concerning operation to avoid instrument damage and personal injury.



The instrument can be equipped with a Class 2 laser. Class 2 lasers shine only within the visible spectrum at an intensity of 1 mW. Looking directly into the laser beam can produce a slight, temporary blinding effect, but does not result in physical injury or damage to the eyes, even when the beam is magnified by optical aids. At any rate, closing the eye lids is encouraged when eye contact is made with the laser beam. Pay attention to possible reflections of the laser beam. The laser functions only to locate and mark surface measurement targets. Do not aim the laser at people or animals.



Pay particular attention to the following safety instructions.



Use in 110/230 V~ electrical systems can result in electrical hazards and personal injury, if not properly protected. All instrument parts supplied by electricity must be covered to prevent physical contact and other hazards at all times.

2 Description

The MI3 sensor series is the next generation of the well-established “MI class” sensor platform. It will be capable of covering a broad range of applications. The MI3 sensor series introduces various network communications, an externally accessible user interface, improved temperature measurement specifications and capabilities at an economic price.

The MI3 series comes with the following highlights:

- Rugged sensing heads survive ambient temperatures to 120°C (248°F) including optimized performance due to ambient temperature calibration across full ambient temperature range
- Special high ambient temperature heads available withstanding up to 180°C (356°F) without any cooling (LTH models)
- Multi head system architecture to allow multiple sensing heads to be multiplexed from a single communication box
- Stand-alone OEM sensing head operation
- Precision high resolution optics up to 100:1
- Up to 10 ms response time
- Alarm status indicator
- Standard USB 2.0 digital interface
- Analog outputs with galvanic isolation
- Alarm relay output
- Serial RS485 communication interface with the protocols: ASCII, Profibus, Modbus¹
- Ethernet communication interface with the protocols: ASCII, http, Profinet IO
- Automatic sensing head identification
- Includes DataTemp[®] Multidrop software for sensor configuration and monitoring
- Field calibration software

¹ Modbus is a registered trademark of Modbus Organization, Inc.

Description

2.1 Overview Comm Boxes





| | MI3COMM metal box | MI3COMMN DIN 3TE | MI3COMM DIN 4TE | MI3COMM... DIN 6TE |
|--|---|---|--|---|
| |  |  |  |  |
| Part number | MI3COMM... | MI3COMMN | MI3COMM | MI3COMM... |
| Spectral Heads¹ | LT, G5, 1M, 2M | LT, G5, 1M, 2M | LT, G5, 1M, 2M | LT, G5, 1M, 2M |
| Head Support by firmware by terminal | 8 heads 1 head | 8 heads 4 heads | 8 heads 4 heads | 8 heads 4 heads |
| Control panel Display Buttons | ✓ ✓ | - - | ✓ ✓ | ✓ ✓ |
| Outputs mA/V TC Relay | 1 1 1 | - - 1 | - - 1 | 4x (optional ...A) - 1 |
| Inputs Emissivity (analog) Emissivity (digital) Ambient Temp. Compensation Trigger/Hold Function Laser Switching | ✓ ✓ ✓ ✓ ✓ | - - - ✓ ✓ | - - - ✓ ✓ | - - - ✓ ✓ |
| Interfaces USB RS485 | Standard Option (...4) | Standard - | Standard Standard | Standard - |
| Protocols ASCII Profibus Modbus Profinet Ethernet | Standard Option (...P1, ...P2) Option (...M) Option (...PN) Option (...E) | Standard - - - - | Standard - - - - | Standard Option (...P) Option (...M) Option (...PN) Option (...E) |

Table 1: Capabilities of Communication Boxes

¹ 1M, 2M spectral heads require box firmware revision 2.11 or higher

3 Technical Data

3.1 Measurement Specification

3.1.1 Sensing Heads

Temperature Range

| | |
|---------------------|-------------------------------|
| LTS02, LTS10, LTH10 | -40 to 600°C (-40 to 1112°F) |
| LTS20, LTF, LTH20 | 0 to 1000°C (32 to 1832°F) |
| G5 | 250 to 1650°C (482 to 3002°F) |
| 2M | 250 to 1400°C (482 to 2552°F) |
| 1M | 500 to 1800°C (932 to 3272°F) |

Spectral Response

| | |
|----|-----------------------|
| LT | 8 to 14 μm |
| G5 | 5 μm |
| 2M | 1.6 μm |
| 1M | 1 μm |

Optical Resolution D:S¹

| | |
|--------|--|
| LTS | 2:1, 10:1, 22:1 typ. (21:1 guaranteed) |
| LTF | 10:1 |
| LTH | 10:1, 22:1 typ. (21:1 guaranteed) |
| G5 | 10:1 |
| 1M, 2M | 100:1 |

SF1 optics: 2 mm spot @ 200 mm distance (0.08 in @ 7.9 in)
SF3 optics: 22 mm spot @ 2200 mm distance (0.87 in @ (8.7 in)

Response Time²

| | |
|---------------------|--------------------|
| LTS (standard), LTH | 130 ms |
| LTF (fast) | 20 ms |
| G5 | 130 ms |
| 1M, 2M | 10 ms ³ |

Accuracy⁴

| | |
|--------|---|
| LT, G5 | \pm (1% of reading or 1°C), whichever is greater \pm 2°C (\pm 4°F) for target temp. < 20°C (68°F) |
| 1M, 2M | \pm (0.5% of reading + 2°C) |

Repeatability

| | |
|--------|--|
| LT, G5 | \pm 0.5% of reading or \pm 0.5°C, whichever is greater |
| 1M, 2M | \pm 0.25% of reading + 1°C |

Temperature Coefficient⁵

| | |
|--------|--|
| LT, G5 | \pm 0.05 K / K or \pm 0.05% of reading / K, whichever is greater |
| 1M, 2M | 0.01% of reading / K |

¹ at 90% energy in minimum and distance 400 mm (15.7 in.)

² 90% response

³ 30 ms – if more than one sensing head drives an analog output of the communication box

⁴ at ambient temperature 23°C \pm 5°C (73°F \pm 9°F), ϵ = 1.0, and calibration geometry

⁵ ambient temperature deviations related to 23°C

Technical Data

3.1.2 Comm Box

Accuracy

| | |
|-------------|---------|
| mA/V output | ± 1°C |
| TC output | ± 1.5°C |

Temperature Resolution

| | |
|-------------|---|
| mA/V Output | ± 0.1°C (± 0.2°F) ¹ / 12 bit, for Comm Box (metal) |
| mA/V Output | ± 0.02°C (± 0.04°F) / 16 bit, for Comm Box (DIN 6TE, analog) |

Temperature Coefficient

| | |
|-------------|--------------|
| mA/V Output | ± 0.02 K / K |
| TC Output | ± 0.05 K / K |

Emissivity

| | |
|------------|----------------|
| All models | 0.100 to 1.100 |
|------------|----------------|

Transmission

| | |
|------------|----------------|
| All models | 0.100 to 1.000 |
|------------|----------------|

3.1.2.1 Comm Box (metal)

Loop Time

| | |
|-------------|-----------------------------------|
| mA/V Output | |
| LTS, G5 | 8 ms |
| LTF, 1M, 2M | 4 ms |
| digital | 18 ms * number of connected heads |

3.1.2.2 Comm Box (DIN)

Loop Time

| | |
|-------------|---|
| digital | |
| LTS, G5 | 8 ms ² * number of connected heads |
| LTF, 1M, 2M | 4 ms ³ * number of connected heads |

¹ for a zoomed temperature span of < 500°C (932°F)

² per bus channel

³ per bus channel

3.2 Optical Charts




Figure 1: Spot Size Charts

Technical Data

3.3 Electrical Specification

For an overview to the capabilities of the communication boxes, see section 2.1 [Overview Comm Boxes](#), page 12.

3.3.1 Comm Box, all models

Voltage Supply 8 to 32 VDC

Power Consumption max. 6 W

Alarm Output

1 potential-free relay output, 48 V / 300 mA

Relay with wear-free contacts (solid state relay) for target temperature or head ambient temperature, electrically isolated from power supply

USB Interface

Version: 2.0

Connector on the board: type Mini-B

3.3.2 Comm Box (metal)

Analog Outputs

Output 1 0 to 5/10 V output for head ambient temperature and object temperature electrically **not** isolated from power supply

TC Thermocouple (type J, K, R, or S)

Output 2 0 to 20 mA (active), or
4 to 20 mA (active), or
0 to 5 V, or
0 to 10 V
electrically **not** isolated from power supply

External Inputs

3 inputs are available useable in different modes:

FTC1-3 Emissivity control: 3 bit digital coded, 0 to V_{SS}

FTC1 Emissivity control: analog, 0 to 5 V_{DC}

FTC2 Ambient temperature compensation analog: 0 to 5 V_{DC}

FTC3 for trigger/hold/laser functions, 0 to V_{SS}

3.3.3 Comm Box (DIN 6TE, analog)

Analog Outputs

Output 1 to 4 0 to 20 mA (active), or
4 to 20 mA (active), or
0 to 5 V, or
0 to 10 V
Each output is galvanically isolated from the other and from power supply!

3.4 Environmental Specification

3.4.1 Sensing Head

Ambient Temperature

| | |
|----------------|--------------------------------------|
| LT, G5 | -10 to 120°C (14 to 248°F) |
| LTH | -10 to 180°C (14 to 356°F) |
| 1M, 2M | 0 to 120°C (32 to 248°F) |
| Laser (1M, 2M) | automatic switch off at 65°C (149°F) |

Storage Temperature

| | |
|------------------|----------------------------|
| LTH | -20 to 180°C (-4 to 356°F) |
| all other models | -20 to 120°C (-4 to 248°F) |

Rating IP65 (NEMA-4) / IEC 60529

Relative Humidity 10% to 95% non-condensing

EMC EN 61326-1:2006

Vibration 11 to 200 Hz, 3 g above 25 Hz operating, 3 axes / IEC 60068-2-6

Shock 50 g, 11 ms, operating, 3 axes / IEC 60068-2-27

Weight

| | |
|--------|----------------|
| LT, G5 | 50 g (1.8 oz) |
| 1M, 2M | 233 g (8.2 oz) |

Material

| | |
|------------------|---|
| Head | Stainless steel |
| Head Cable | |
| LTH | Teflon® |
| all other models | PUR (Polyurethane), Halogen free, Silicone free |



Teflon develops poisonous gasses when it comes into contact with flames!

3.4.2 Comm Box (metal)

Ambient Temperature -10 to 65°C (14 to 149°F)

Storage Temperature -20 to 85°C (-4 to 185°F)

Rating IP65 (NEMA-4) / IEC 60529

Relative Humidity 10% to 95% non-condensing

EMC EN 61326-1:2006

Vibration 11 to 200 Hz, 3 g above 25 Hz operating, 3 axes / IEC 60068-2-6

Shock 50 g, 11 ms, operating, 3 axes / IEC 60068-2-27

Weight 370 g (13 oz)

Material die-cast zinc enclosure

Technical Data

3.4.3 Comm Box (DIN)

| | |
|---------------------|---|
| Ambient Temperature | -10 to 65°C (14 to 149°F) |
| Storage Temperature | -20 to 85°C (-4 to 185°F) |
| Relative Humidity | 10% to 95% non-condensing |
| EMC | EN 61326-1:2006 |
| Vibration | 11 to 200 Hz, 3 g above 25 Hz operating, 3 axes / IEC 60068-2-6 |
| Shock | 50 g, 11 ms, operating, 3 axes / IEC 60068-2-27 |
| Weight | 125 g (4.4 oz) |
| Material | molded plastic |

3.4.4 LTH Electronics

| | |
|---------------------|---------------------------|
| Ambient Temperature | -10 to 65°C (14 to 149°F) |
| Storage Temperature | -20 to 85°C (-4 to 185°F) |
| Rating | IP65 (NEMA-4) / IEC 60529 |

3.5 Dimensions

3.5.1 Sensing Head LT, G5




Figure 2: Dimensions of LT, G5 Sensing Heads

3.5.2 Sensing Head LTH




Figure 3: Dimensions of LTH Sensing Head with separated Electronics

3.5.3 Sensing Head 1M, 2M




Figure 4: Dimensions of 1M, 2M Sensing Heads

3.5.4 Comm Box (metal)

The box is equipped with three cable feed-through ports – two with IP65 compatible sealing glands, a third sealing gland comes for boxes with fieldbus communications (RS485, Profibus etc.). Boxes without fieldbus have a plugged expansion feed-through port instead (M12x1.5 thread).





Figure 5: Dimensions of Communication Box

Technical Data

3.5.5 Comm Box (DIN)

The boxes come in a standard DIN rail size in accordance to EN 50022-35x7.5 (DIN 43880).



| Width | MI3COMMN | MI3COMM | MI3COMM... |
|-------|------------------------------|------------------------------|-------------------------------|
| X | DIN 3TE: 53.6 mm (2.1 in) | DIN 4TE: 71.6 mm (2.8 in) | DIN 6TE: 107.6 mm (4.2 in) |

Figure 6: Dimensions for Comm Boxes (DIN)

3.6 Scope of Delivery

3.6.1 Sensing Head

- Sensing head with 1 m (3 ft.) cable
- Laser (1M, 2M heads only)
- Mounting nut

3.6.2 Comm Box

- Communication box
- Software DVD
- Quickstart guide

4 Basics

4.1 Measurement of Infrared Temperature

All surfaces emit infrared radiation. The intensity of this infrared radiation changes according to the temperature of the object. Depending on the material and surface properties, the emitted radiation lies in a wavelength spectrum of approximately 1 to 20 μm . The intensity of the infrared radiation (heat radiation) is dependent on the material. For many substances, this material-dependent constant is known. This constant is referred to as the "emissivity value".

Infrared thermometers are optical-electronic sensors. These sensors are sensitive to the emitted radiation. Infrared thermometers are made up of a lens, a spectral filter, a sensor, and an electronic signal processing unit. The task of the spectral filter is to select the wavelength spectrum of interest. The sensor converts the infrared radiation into an electrical signal. The signal processing electronics analyze the electrical signal and convert it into a temperature measurement. As the intensity of the emitted infrared radiation is dependent on the material, the required emissivity can be selected on the sensor.

The biggest advantage of the infrared thermometer is its ability to measure temperature without touching an object. Consequently, surface temperatures of moving or hard to reach objects can easily be measured.

4.2 Emissivity of Target Object

To determine the emissivity of the target object see section 19.1 [Determination of Emissivity](#), page 115. If emissivity is low, measured results could be falsified by interfering infrared radiation from background objects (such as heating systems, flames, fireclay bricks, etc. located close beside or behind the target object). This type of problem can occur when measuring reflective surfaces and very thin materials, such as plastic film and glass.

This measurement error can be reduced to a minimum, if particular care is taken during installation and the sensing head is shielded from these reflecting radiation sources.

4.3 Ambient Temperature

The sensing head is suited for the ambient temperatures up to 120°C (248°F) for the standard heads and up to 180°C (356°F) for the LTH heads. The sensing head can operate in ambient temperatures up to 200°C (392°F) with the air-cooling accessory.

4.4 Atmospheric Quality

If the lens gets dirty, infrared energy will be blocked and the instrument will not measure accurately. It is good practice to always keep the lens clean. The Air Purge Jacket helps keep contaminants from building up on the lens. If you use air purging, make sure a filtered air supply with clean dry air at the correct air pressure is installed before proceeding with the sensor installation.

4.5 Electrical Interference

To minimize electrical or electromagnetic interference or "noise", please be aware of the following:

- Mount the unit as far away as possible from potential sources of electrical interference, such as motorized equipment, which can produce large step load changes.
- Use shielded wire for all input and output connections.
- To avoid current equalizations, make sure that a sufficient potential equalization is realized between the sensing head and metal housing of the communication box.

Basics

- To avoid ground loops, make sure that only **one point** is earth grounded, either via the sensing head, the Comm Box, or power.

Please note that:

- The metal housings of the sensing head and the MI3 communication box are electrically connected to the shield of the head cable.
- All inputs and outputs (except the alarm output and the outputs of the Comm Box (DIN 6TE, analog)) use the same ground and are electrically connected to the power supply.




Figure 7: Shield Run for Comm Box (metal)




Figure 8: Shield Run for Comm Box (DIN)




Figure 9: Only one point is earth grounded either via sensing head, via Comm Box, or via Power

5 Installation

5.1 Positioning

Sensor location depends on the application. Before deciding on a location, you need to be aware of the ambient temperature of the location, the atmospheric quality of the location, and the possible electromagnetic interference in that location. If you plan to use air purging, you need to have an air connection available. Wiring and conduit runs must be considered, including computer wiring and connections, if used.

5.1.1 Distance to Object

The desired spot size on the target will determine the maximum measurement distance. To avoid erroneous readings, the target spot size must completely fill the entire field of view of the sensor. Consequently, the sensor must be positioned so the field of view is the same as or smaller than the desired target size. For a list indicating the available optics, see section 3.2 [Optical Charts](#), page 15.

The actual spot size for any distance can be calculated by using the following formula. Divide the distance D by your model's D:S number. For example, for a unit with D:S = 10:1, if the sensor is 400 mm (15.7 in.) from the target, divide 400 by 10 (15.7 by 10), which gives you a target spot size of approximately 40 mm (1.57 in.).




Figure 10: Proper Sensor Placement

5.2 Installation Schemes

5.2.1 Comm Box (metal)

The basic stand-alone configuration consists of one sensing head interfaced to one metallic communications box. The sensing head provides all IR measurement functionality. The communications box provides an externally accessible user interface and display, advanced signal processing capability, field wiring terminations and fieldbus functionality with optional RS485 communication interface.

Installation




Figure 11: Single Head Configuration with Comm Box

To increase the number of supported sensing heads, you can use a dedicated accessory, see section 10.1.1 [Multi-Channel Box](#), page 55.

5.2.2 Comm Box (DIN)

The multiple sensing head configuration consists of a modular communication box provided in a DIN rail mountable plastic enclosure for supporting 4 sensing heads simultaneously. The DIN rail communication box provides an externally accessible user interface. The terminal strip connectors are used to simplify the field wiring.




Figure 12: Multiple Head Configuration with DIN Rail Comm Box

5.3 Wiring, Head Cable

The user has to install the sensor cable on the communication box. It may be shortened, if necessary, but keep a minimal length of 20 cm (7.9 in.).

Do not bend the sensing head cable tighter than a radius of 25 mm (1 in.) for the standard heads (PUR cable) and 15 mm (0.6 in.) for the high ambient temperature heads (Teflon cable) respectively!



The total sensing head cable length for all networked sensing heads must not exceed 30 m (98 ft) for MI3 and 2x30 m (2x98 ft) for MI3M!



Do not add a third party cable to extend the length of the sensing head cable!

5.3.1 Comm Box (metal)

1. Cut about 40 mm (1.5 in) of the cable sheath from the end of the sensing head cable ⑦.
Caution: Do not cut into the shield!
2. Cut the shield ⑤ so about 5 mm (0.2 in) remains exposed from under the cable sheath. Separate the shield and spread the strands out.
3. Strip 3 mm (0.12 in) of insulation from the wires ⑥!
4. Open the communication box by removing the four Phillips head screws and pulling off the lid. Unscrew the pressure screw ①, and remove the first metal washer ④, the rubber washer ③, and the second and the third metal washers ④.
5. Put the following on the cable: the pressure screw ①, the first metal washer ④, the rubber washer ③ and the second metal washers ④, see the following figure.
6. Spread the cable shield ⑤ and then slip the third metal washer ④ onto the cable. Note that the shield must make good contact to both metal washers.
7. Slip the wires ⑥ into the communication box far enough to connect to the terminal.
8. Screw the pressure screw ① into the communication box. Tighten snugly. Do not over tighten.
9. Connect the wires ⑥ to the terminal on the printed circuit board.




Figure 13: Sensing Head Cable to the Comm Box

5.3.2 Comm Box (DIN)

The wiring of the sensing head cable is color coded, see section 5.4.3 [Comm Box \(DIN 4 TE\)](#), page 28.

5.4 Wiring, Terminal

You need to connect the power supply and possibly the signal input/output wires. Use only cable with outside diameter from 4 to 6 mm (0.16 to 0.24 in), wire size: 0.14 to 0.75 mm² (AWG 19 to 26).

Installation



The cable must include shielded wires. It should not be used as a strain relief!

5.4.1 Comm Box (metal)




Figure 14: Terminal Wiring for the Comm Box

5.4.2 Comm Box (DIN 3TE)




Figure 15: Terminal Wiring for the Comm Box DIN 3TE

Installation

5.4.3 Comm Box (DIN 4 TE)




Figure 16: Terminal Wiring for the Comm Box DIN 4 TE

5.4.4 Comm Box (DIN 6 TE)




Figure 17: Terminal Wiring for the Comm Box DIN 6 TE for Profibus and Modbus

Installation




Figure 18: Terminal Wiring for the Comm Box DIN 6 TE for Profinet and Modbus




Figure 19: Terminal Wiring for the Comm Box DIN 6 TE, analog

Installation

5.4.5 EMI Resistance for Comm Box (DIN)



To maintain EMI compliance to CE standards the attached Ferrite cores need to be placed on all wires! Make sure that the cable shields will be connected to the terminal pin <Shield>!




Figure 20: Mounting of Shield Wire and Ferrite Core

5.5 Power On Procedure

To power the system, the following procedures are required.

5.5.1 One Head System

1. Disconnect power to the box.
2. Connect the wires for the head to the box terminal.
3. Power the box.
4. The box now assigns address 1 to the head.

5.5.2 Multiple Heads – Random Address Assignment

1. Disconnect power to the box.
2. Connect the wires for all heads to the box terminal.
3. Power the box.
4. The box automatically assigns a unique address to each of the heads – the mapping of physical head and head address is randomly.

5.5.3 Multiple Heads – User Controlled Address Assignment

1. Disconnect power to the box.
2. Connect the wires for the first head to the box terminal.
3. Power the box.
4. The box now assigns address 1 to the first head.
5. Follow the instructions 1 to 4 to add the next head. With each new head detected, the box increases the head address by 1.



The head address may be changed later by the user under the dedicated head page. See section 8.2 [<Head> Page](#), page 47.

Installation

5.6 USB

The USB interface comes with each box (USB connector, Mini-B). Connect a single unit to a USB computer port by using an appropriate USB cable.




Figure 21: USB Connection via the Comm Box (metal)





Figure 22: USB Connection via the Comm Box (DIN Rail)

Consider the following sequence for the installation:

1. Disconnect/reconnect the USB interface cable to the computer!
2. Ignore the Windows Wizard <Found New Hardware>!
3. Navigate manually to the dedicated USB driver <RaytekMIcomport.inf> on the support media and execute it.



It is strongly recommended to check the correct driver installation under the Windows Operating System <Start> <Settings> <Control Panel> <System> <Hardware> <Device Manager> <Ports (COM & LPT)>. Go there also to get the virtual COM port number for communicating with the DTMD Software.



Driver correctly installed!

COM port number
for DTMD Software!

Installation

5.7 Fieldbus



It is strongly recommended to avoid any simultaneous communication via USB and fieldbus!

5.7.1 Addressing



Each slave in the network must have a unique address and must run at the same baud rate!

For setting the fieldbus configurations through the control panel, see section 8.3 [<Box Setup> Page](#), page 48.

5.7.2 RS485 based Installations

The recommended way to add more devices into a network is connecting each device in series to the next in a linear topology (daisy chain).

Use only one power supply for all boxes in the network to avoid ground loops!



It is strongly recommended to use shielded and pair twisted cables (e.g. CAT.5)!



Make sure the network line is terminated!





Figure 23: Network in Linear Topology (daisy chain)

6 Outputs

For the outputs the following groupings (setups) are possible:

| Output | Setup 1 | Setup 2 | Setup 3 | Setup 4 | Setup 5 |
|--------|--------------------------|-------------------------|--------------------------|-------------------------|-----------------------|
| OUT1 | head temperature V | head temperature V | object temperature V | object temperature V | – |
| OUT2 | object temperature mA | object temperature V | object temperature mA | object temperature V | head temperature V |
| TC | – | – | – | – | object temperature |

6.1 Analog Output OUT1

| | | |
|---|------------------|---|
|  | Comm Box: | metal |
| | Source: | object temperature / head ambient temperature |
| | Signal: | 0 to 5/10 V |
| | Terminal: | OUT1, GND |


This output can be configured for the object or the head ambient temperature. E.g. the output range for the head ambient temperature is 0 to 5 VDC corresponding to 0 to 500°C (32 to 932°F). The minimum load impedance for the voltage output must be 10 kΩ.

The output is short circuit resistant.



The outputs <OUT1> and <TC> are not available at the same time!

6.2 Analog Output OUT2


| | | |
|---|------------------|---|
|  | Comm Box: | metal |
| | Source: | object temperature / head ambient temperature |
| | Signal: | 0/4 to 20 mA or 0 to 5/10 V |
| | Terminal: | OUT2, GND |

The signal output can be configured as either current or voltage output. The minimum load impedance for the voltage output must be 10 kΩ. The maximum current loop impedance for the mA output is 500 Ω.

The output is short circuit resistant.

Outputs

6.3 Analog Outputs OUT1 - OUT4

| | | |
|---|------------------|--|
|  | Comm Box: | DIN 6TE, analog (4 channels) |
| | Source: | object temperature / head ambient temperature |
| | Signal: | 0/4 to 20 mA or 0 to 5/10 V |
| | Terminal: | I ₁₋₄ , U ₁₋₄ , GND ₁₋₄ |

Each signal output can be configured as either current or voltage output, whereby each sensing head can be assigned to each output. The minimum load impedance for the voltage output must be 10 kΩ. The maximum current loop impedance for the mA output is 500 Ω. All outputs are short circuit resistant.




Each output is galvanically isolated from the other and from the power supply!



Powering the communication box (DIN, 6TE), analog, only via USB will disable all analog outputs! The configuration of the box under the control panel is furthermore possible. Under the DataTemp Multidrop you can only drive and configure the analog outputs by using an additional external power supply!

6.4 Alarm Output RELAY

| | | |
|---|------------------|---|
|  | Comm Box: | all models |
| | Source: | object temperature / head ambient temperature |
| | Signal: | potential-free contacts |
| | Terminal: | RELAY, RELAY |

The alarm output is controlled by the target object temperature or the head ambient temperature. In case of an alarm, the output switches the potential free contacts from a solid state relay. The maximum load for this output is 48 V / 300 mA.

If a spike voltage exceeding the absolute maximum rated value is generated between the output terminals, insert a clamping diode in parallel to the inductive load as shown in the following circuit diagram to limit the spike voltage.





Figure 24: Spike Voltage Limitation for the Alarm Relay

6.5 Thermocouple Output TC

| | | |
|---|------------------|-----------------------|
|  | Comm Box: | metal |
| | Source: | object temperature |
| | Signal: | TCJ, TCK, TCR, or TCS |
| | Terminal: | TC, GND |

This output can be configured as thermocouple output type J, K, R, or S. For that output, you must install a dedicated compensation cable. The output impedance is 20 Ω . The output is short circuit resistant.



The outputs <OUT1> and <TC> are not available at the same time!

Inputs

7 Inputs

Three external inputs FTC1, FTC2, and FTC3 are used for the external control of the unit.




You cannot enable the input functions through the control panel!

| | FTC1 | FTC2 | FTC3 |
|---|------|------|------|
| Emissivity (analog control) | x | | |
| Emissivity (digital control) | x | x | x |
| Ambient Background Temperature Compensation | | x | |
| Trigger/Hold Function | | | x |
| Laser Switching | | | x |

Table 2: Overview for FTC Inputs

7.1 Emissivity (analog)

| | | |
|---|------------------|-----------------------------|
|  | Function: | emissivity (analog control) |
| | Signal: | 0 to 5 Vdc |
| | Terminal: | FTC1, GND |

The FTC1 input can be configured to accept an analog voltage signal (0 to 5 VDC) to provide real time emissivity setting. Each input can support one head. The following table shows the relationship between input voltage and emissivity:

| | | | | | |
|-------------------|-----|-----|-----|-----|-----|
| U in V | 0.0 | 0.5 | ... | 4.5 | 5.0 |
| Emissivity | 0.1 | 0.2 | ... | 1.0 | 1.1 |

Table 3: Ratio between Analog Input Voltage and Emissivity

Example:

This process requires setting the emissivity:

- for product 1: 0.90
- for product 2: 0.40

Following the example below, the operator needs only to switch to position “product 1” or “product 2”.





Figure 25: Adjustment of Emissivity at FTC Input (Example)

7.2 Emissivity (digital)

| | | |
|---|------------------|------------------------------|
|  | Function: | emissivity (digital control) |
| | Signal: | digital low/high |
| | Terminal: | FTC1-3, GND |

The box electronics contains a table with 8 pre-installed settings for emissivity. To activate these emissivity settings, you need to have the inputs FTC1, FTC2, and FTC3 connected. According to the voltage level on the FTC inputs, one of the table entries will be activated.

0 = Low signal (0 V)

1 = High signal (from 5 V to V_{DC})


A non-wired input is considered as not defined!

| Table entry | Emissivity (Examples) | FTC3 | FTC2 | FTC1 |
|-------------|-----------------------|------|------|------|
| 0 | 1.100 | 0 | 0 | 0 |
| 1 | 0.500 | 0 | 0 | 1 |
| 2 | 0.600 | 0 | 1 | 0 |
| 3 | 0.700 | 0 | 1 | 1 |
| 4 | 0.800 | 1 | 0 | 0 |
| 5 | 0.970 | 1 | 0 | 1 |
| 6 | 1.000 | 1 | 1 | 0 |
| 7 | 0.950 | 1 | 1 | 1 |

Figure 26: Digital Selection of Emissivity with FTC Inputs

The values in the table cannot be changed through the control panel.

7.3 Ambient Temperature Compensation

| | | |
|---|------------------|----------------------------------|
|  | Function: | Ambient Temperature Compensation |
| | Signal: | 0 to 5 V _{DC} |
| | Terminal: | FTC2, GND |

Inputs

The sensor is capable of improving the accuracy of target temperature measurements by taking into account the ambient or background temperature. This feature is useful when the target emissivity is below 1.0 and the background temperature is significantly hotter than the target temperature. For instance, the higher temperature of a furnace wall could lead to hotter temperatures being measured especially for low emissivity targets.

Ambient background temperature compensation allows for the impact of reflected radiation in accordance with the reflective behavior of the target. Due to the surface structure of the target, some amount of ambient radiation will be reflected and therefore, added to the thermal radiation that is collected by the sensor. The ambient background temperature compensation adjusts the final result by subtracting the amount of ambient radiation measured from the sum of thermal radiation the sensor is exposed to.



The ambient background temperature compensation should always be activated in case of low emissivity targets measured in hot environments or when heat sources are near the target!

Three possibilities for ambient background temperature compensation are available:

- The **internal sensing head temperature** is utilized for compensation assuming that the ambient background temperature is more or less represented by the internal sensing head temperature. This is the default setting.
- If the background ambient temperature is known and constant, the user may give the known ambient temperature as a **constant temperature value**.
- Ambient background temperature compensation from a **second temperature sensor** (infrared or contact temperature sensor) ensures extremely accurate results. For example, a second IR sensor, configured to provide a 0 to 5 volt output scaled for the same temperature range as the target can be connected to input FTC2 to provide real-time ambient background compensation.




Figure 27: Principle of Ambient Background Temperature Compensation

7.4 Trigger/Hold

| | | |
|---|------------------|------------------|
|  | Function: | Trigger/Hold |
| | Signal: | digital low/high |
| | Terminal: | FTC3, GND |

The FTC3 input can be used as an external trigger functioning as “Trigger” or “Hold”. All sensing heads are effected by the FTC3 input at the same time.




Figure 28: Wiring of FTC3 as Trigger/Hold

Trigger: A logical low signal at the input FTC3 will reset the peak or valley hold function. As long as the input is kept at logical low level, the software will transfer the actual object temperatures toward the output. At the next logical high level, the hold function will be restarted.




Figure 29: FTC for Resetting the Peak Hold Function

Hold: This mode acts as an externally generated hold function. A transition at the input FTC3 from logical high level toward logical low level will transfer the current temperature toward the output. This temperature will be written to the output until a new transition from high to low occurs at the input FTC3.

Inputs





Figure 30: FTC3 for Holding the Output Temperature

7.5 Laser Switching

| | | |
|---|------------------|------------------------|
|  | Function: | Laser switching on/off |
| | Signal: | digital low/high |
| | Terminal: | FTC3, GND |

The FTC3 input can also be used as an external trigger to switch the laser (only available for selected sensing head models). A transition at the input from logical high level toward logical low level will switch the laser. All sensing heads are effected by the FTC3 input at the same time.




Figure 31: Wiring of FTC3 as Laser Switching

8 Operation

Once you have the sensor positioned and connected properly, the system is ready for continuous operation.

The control panel is accessible on the outside of the box. Push buttons provide positive tactile feedback to the user. User interface includes a backlit LCD, displaying sensor set up parameters and temperature outputs. Alternatively, the operation of the sensor can be done by means of the software that came with your sensor.

8.1 Control Panel

The sensor system is equipped with a control panel integrated in the box lid, which has setting/controlling buttons and an LCD display.




Figure 32: Control Panel for the Comm Box (metal)



Figure 33: Control Panel for the Comm Box (DIN)




Figure 34: Elements of the LCD Display

Operation

The head number is shown only if two or more sensing heads are connected to the communication box.

| Symbol/Message | Meaning | Remark |
|--|--|--|
| AVG | Average | |
| PH | Peak Hold | |
| VH | Valley Hold | |
| HOLD | Trigger set to HOLD function | |
| APH | Advanced Peak Hold | Software controlled |
| APHA | Advanced Peak Hold with Averaging | Software controlled |
| AVH | Advanced Valley Hold | Software controlled |
| AVHA | Advanced Valley Hold with Averaging | Software controlled |
| <Power Fault> & alarm indicator are blinking | Power via USB not sufficient to drive all analog outputs of Communication box (DIN, 6TE), analog | Configuration of the box is possible but outputs are set to disabled |

Table 4: Symbols and Messages in the Display

Pushing the keys of the control panel will cause the following actions:




enters the menu or save parameters



enters the next page

No action for 10 s forces the unit to leave the menu without saving of parameters.



* not available for all models
4 output channels for Comm Box DIN 6 TE, analog

8.2 <Head> Page

- <Tambient> current head ambient temperature
- <Emissivity> changes the emissivity value for the selected head. The emissivity is a calculated ratio of infrared energy emitted by an object to the energy emitted by a blackbody at the same temperature (a perfect radiator has an emissivity of 1.00). For information on determining an unknown emissivity and for sample emissivities, see section 19.2 [Typical Emissivity Values](#), page 115.
- <Transmiss.> changes the transmission value when using protective windows. For example, if a protective window is used with the sensor, set the transmission to the appropriate value.

Operation





| | |
|------------------------------|---|
| <Laser> | handles the laser in the following modes: <off> switches the laser off <flash> forces the laser to blink at 8 Hz <on> switches the laser permanently on <external> switches the laser via external input FTC3 An activated laser will be switched off automatically after 10 minutes. The laser is available for 1M and 2M heads only. The laser can be activated at the same time for maximal 4 heads. |
| <Average> AVG | signal post processing set to averaging, parameter given in seconds. Once <Average> is set above 0 s, it automatically activates. Note that other hold functions (like Peak Hold or Valley Hold) cannot be used concurrently. Value range: 0.0 to 998.9 sec, ∞ See section 8.5.1 Averaging , page 50. |
| <Peak Hold> PH | signal post processing set to Peak Hold, parameter given in seconds. Once <Peak Hold> is set above 0 s, it automatically activates. Note that other hold functions (like Valley Hold or Averaging) cannot be used concurrently. Value range: 0.0 to 998.9 sec, ∞ See section 8.5.2 Peak Hold , page 51. |
| <Val. Hold> VH | signal post processing set to Valley Hold, parameter given in seconds. Once <Valley Hold> is set above 0 s, it automatically activates. Note that other hold functions (like Peak Hold or Averaging) cannot be used concurrently. Value range: 0.0 to 998.9 sec, ∞ See section 8.5.3 Valley Hold , page 51. |
| <Trigger> | defines the trigger mode for the selected head: <trig>: ... to reset the peak or valley hold function <hold>: activates the hold function See section 7.4 Trigger/Hold , page 43. |
| <Alarm Mode> | defines the alarm mode for the selected head: <Tobj>: object temperature as alarm source <Tamb>: head ambient temperature as alarm source |
| <Set Point> | defines a temperature threshold for an alarm |
| <Lo Limit> | low end of temperature measurement range (read only) |
| <Hi Limit> | high end of temperature measurement range (read only) |
| <Fact. default> | sets the selected head back to factory default. The factory default values are to be found in section 18.9 Command Set , page 108. |
| <Type> | provides the head model, e.g. MI3LT |
| <SN> | provides the serial number for the selected head and allows to reassign a new head address |

8.3 <Box Setup> Page

| | |
|---------------------------|---|
| <Relay Mode> | defines the switching behavior for the box internal alarm relay: <normally open>: open contact in non-alarm status <normally closed>: closed contact in non-alarm status <permanently OFF>: permanently open contacts <permanently ON>: permanently closed contacts |
|---------------------------|---|

| | |
|-------------------|---|
| <OUT1 Mode> | defines the mode for the analog output: <TCJ>, <TCK>, <TCR>, <TCS> <0-5V> <0-10V> <disable> output goes to high-resistance |
| <OUT1 Source> | assigns the selected head to the analog output: <#1>, <#2>, ..., <Head _{max} > |
| <OUT1 Value> | defines the basis for the output value: <Tobject>: object temperature to be output <Tambient>: head ambient temperature to be output |
| <OUT1 low temp.> | defines the temperature for the low end of the analog output range (scaling) |
| <OUT1 high temp.> | defines the temperature for the high end of the analog output range (scaling) |
| <OUT2 Mode> | defines the mode for the analog output 2: <0-20mA> <4-20mA> <0-5V> <0-10V> <disable> output goes to high-resistance |
| <OUT2 Source> | assigns the selected head to the analog output: <#1>, <#2>, ..., <Head _{max} > |
| <OUT2 Value> | defines the basis for the output value: <Tobject>: object temperature to be output <Tambient>: head ambient temperature to be output |
| <OUT2 low temp.> | defines the temperature for the low end of the analog output range (scaling) |
| <OUT2 high temp.> | defines the temperature for the high end of the analog output range (scaling) |
| <Interface> | RS485/Profibus/Modbus: <address>: unique address of the box in the network. <baudrate>: baud rate for the box. Each device in the network must be set to the same baud rate. The baud rate for Profibus is automatically negotiated between master and slave. Profinet/Ethernet: <DHCP>: on/off – network protocol to configure a device in a network. DHCP is switched off for Profinet communications. <IP address>: a unique address of the box in the network, only changeable at DHCP = off <SubNetMask>: the subnet mask defines the interpretation of the IP address, only changeable at DHCP = off <Gateway>: a gateway connects two subnets at different subnet addresses, only changeable at DHCP = off <Port>: communication port, only changeable at DHCP = off, read-only for Profinet communications <MAC>: MAC address, read-only |
| <Factory default> | sets the box back to factory default. The factory default values are to be found in section 18.9 Command Set , page 108. |

Operation

- <Temperature Unit>** the temperature unit can be set to °C or °F. Note that this setting influences the digital interfaces like RS485 for both object and head ambient temperature.
- <Key Enter Lock>** the box has a user interface lockout feature that keeps the box from being accidentally changed from the control panel (locked by default under DataTemp Multidrop Software and Profinet communications). This lockout mode denies access to the  button to avoid the saving of adjustable parameters. The unit can be unlocked by pressing the  button and the  button simultaneously for 3 seconds or alternatively by pressing the  button for 5 seconds.
- <Display Backlight>** defines the switching behavior for the display:
<ON>: switches the backlight on
<OFF>: switches the backlight off
<60sec.OFF>: switches the backlight off after the giving time
To preserve the display's longevity, the backlight should be turned off in case of not using it!

8.4 <Box Info> Page

- <Serial No.>**: serial number of the box.
<Rev>: firmware revision
Tbox: current box ambient temperature

8.5 Post Processing

8.5.1 Averaging

Averaging is used to smooth the output signal. The signal is smoothed depending on the defined time basis. The output signal tracks the detector signal with significant time delay but noise and short peaks are damped. Use a longer average time for more accurate damping behavior. The average time is the amount of time the output signal needs to reach 90% magnitude of an object temperature jump.




Figure 35: Averaging

A low level input (GND) at external input FTC3 will promptly interrupt the averaging and will start the calculation again.

Attention: The disadvantage of averaging is the time delay of the output signal. If the temperature jumps at the input (hot object), the output signal reaches only 90% magnitude of the actual object temperature after the defined average time.

8.5.2 Peak Hold

The output signal follows the object temperature until a maximum is reached. The output will „hold“ the maximum value for the selected duration of the hold time. Once the hold time is exceeded, the peak hold function will reset and the output will resume tracking the object temperature until a new peak is reached. The range for the hold time is 0.1 to 998.9 s.




Figure 36: Peak Hold

A defined hold time of 999 s (symbol “∞” in the display) will put the device into continuous peak detection mode.

A low level input (GND) at external input FTC3 will promptly interrupt the hold time and will start the maximum detection again.

8.5.3 Valley Hold

The output signal follows the object temperature until a minimum is reached. The output will „hold“ the minimum value for the selected duration of the hold time. Once the hold time is exceeded, the valley hold function will reset and the output will resume tracking the object temperature until a new valley is reached. The range for the hold time is 0.1 to 998.9 s

Operation




Figure 37: Valley Hold

A defined hold time of 999 s (symbol “∞” in the display) will put the device into continuous valley detection mode.

A low level input (GND) at external input FTC3 will promptly interrupt the hold time and will start the minimum detection again.

8.5.4 Advanced Peak Hold

This function searches the sensor signal for a local maximum (peak) and writes this value to the output until a new local maximum is found. Before the algorithm restarts its search for a local maximum, the object temperature has to drop below a predefined threshold. If the object temperature rises above the held value, which has been written to the output so far, the output signal follows the object temperature again. If the algorithm detects a local maximum while the object temperature is currently below the predefined threshold, the output signal jumps to the new maximum temperature of this local maximum. Once the actual temperature has passed a maximum above a certain magnitude, a new local maximum is found. This magnitude is called hysteresis.




Figure 38: Advanced Peak Hold

The advanced peak hold function is only adjustable by means of the DataTemp Multidrop Software.

8.5.5 Advanced Valley Hold

This function works similar to the advanced peak hold function, except that it will search the signal for a local minimum.

8.5.6 Advanced Peak Hold with Averaging

The output signal delivered by the advanced peak hold functions tends to jump up and down. This is due to the fact, that only maximum points of the otherwise homogenous trace will be shown. The user may combine the functionality of the peak hold function with the averaging function by choosing an average time, thus, smoothing the output signal for convenient tracing.




Figure 39: Advanced Peak Hold with Averaging

The advanced peak hold function with averaging is only adjustable by means of the DataTemp Multidrop Software.

8.5.7 Advanced Valley Hold with Averaging

This function works similar to the advanced peak hold function with averaging, except it will search the signal for a local minimum.

Options

9 Options

Options are items that are factory installed and must be specified at time of order. The following are available:

- Longer head cables in the lengths:
 - 3 m / 9.8 ft. (...CB3)
 - 8 m / 262 ft. (...CB8)
 - 15 m / 49 ft. (...CB15)
 - 30 m / 98 ft. (...CB30)
- Network communication interfaces:
 - RS485, Profibus DP, Modbus RTU, Profinet IO, EthernetAll available models are listed under section 2.1 [Overview Comm Boxes](#), page 12.

10 Accessories

A full range of accessories for various applications and industrial environments are available. Accessories include items that may be ordered at any time and added on-site.

10.1 Accessories (all models)

- [Multi-Channel Box](#) (XXXMI3CONNBOX)
- [USB/RS485 Adapter](#) for boxes with RS485 interface (XXXUSB485)

10.1.1 Multi-Channel Box

The Multi-Channel Box can be used for all communication boxes. The box includes 8 sets of field wiring terminals wired in parallel to one 5 m (16 ft) cable set to connect to the communication box.




Figure 40: Multiple Head Configuration with Comm Box




Figure 41: Dimensions

Accessories




Figure 42: Wiring Diagram for 8 Heads

Please note the correct mounting of the cable shield requires a strong metallic contact to the grommet.




Figure 43: Correct Mounting of the Cable Shield



The total sensing head cable length for all networked sensing heads must not exceed 30 m/98 ft (for MI3) and 2x30 m/2x98 ft (for MI3M)!

10.1.2 USB/RS485 Adapter

The USB/RS485 adapter is self-powering via the USB connection.



Figure 44: USB/RS485 Adapter (XXXUSB485)

Accessories




Figure 45: Wiring the RS485 Interface of the Box (left) and USB/RS485 Adapter (right)

10.2 Accessories (LT, G5 Heads)

- [Adjustable Mounting Bracket](#) (XXXMIACAB)
- [Fixed Mounting Bracket](#) (XXXMIACFB)
- Sensing head mounting nut (XXXMIACMN)
- [Air Purge Jacket](#) (XXXMIACAJ)
- [Air Cooling System](#) with 0.8 m (2.6 ft.) air hose (XXXMIACCJ) or with 2.8 m (9.2 ft.) air hose (XXXMIACCJ1)
- [Right Angle Mirror](#) (XXXMIACRAJ, XXXMIACRAJ1)
- [Protective Windows](#)
- [Close Focus Lens](#) (XXXMI3ACCFL)




Figure 46: Standard Accessories for LT, G5 Heads

Accessories

10.2.1 Adjustable Mounting Bracket




Figure 47: Adjustable Mounting Bracket (XXXMIACAB)

10.2.2 Fixed Mounting Bracket




Figure 48: Fixed Mounting Bracket (XXXMIACFB)

10.2.3 Air Purge Jacket



For LTH sensing heads, the Air Purge Jacket is only available pre-mounted from the factory (XXXMIACAJI)!

The air purge jacket is used to keep dust, moisture, airborne particles, and vapors away from the sensing head. Clean, oil free air is recommended. The air purge jacket withstands ambient temperatures up to 180°C (356°F) and has limited use for cooling purposes. The recommended air flow rate is 30 to 60 l / min (0.5 to 1 cfm). The max. pressure is 5 bar (73 PSI).

Accessories




Figure 49: Air Purge Jacket (XXXMIACAJ)




Figure 50: Mounting the Air Purge Jacket

1. Remove the sensor ① and cable from the communication box by disconnecting the wires from the terminal.
2. Open the Air Purge Jacket ③ ④ and screw the white plastic fitting ② onto the sensor up to the end of the threads. Do not over-tighten!
3. Slip the cable ⑥ through the backside ④ of the jacket.
4. Close the Air Purge Jacket ③ ④, reconnect the wires to the communication box and apply the mounting nut ⑤.

10.2.4 Air Cooling System



The Air Cooling System can not be combined with LTH heads!

The sensing head can operate in ambient temperatures up to 200°C (392°F) with the air-cooling system. The air-cooling system comes with a T-adaptor including 0.8 m / 31.5 in (optional: 2.8 m / 110 in) air hose and insulation. The T-adaptor allows the air-cooling hose to be installed without interrupting the connections to the box.

The air-cooling jacket may be combined with the right angle mirror.




Figure 51: Air Cooling System (XXXMIACC)




Figure 52: Connecting the T-Adapter

Accessories




Figure 53: Maximum Ambient Temperature is dependent on Air Flow and Hose Length

Note: "Hose Length" is the length of the hose exposed to high ambient temperature (not the overall length of the hose).




Figure 54: Air Cooling System: Purging Jacket and T-Adapter

The Air Cooling System consists of:

- ① sensing head
- ② inner plastic fitting (air purge jacket)
- ③ front part of the air-purging jacket
- ④ back part of the air-purging jacket
- ⑤ mounting nut
- ⑥ preinstalled cable between sensor and box, leading through the T-adapter
- ⑦ hose connecting nut
- ⑧ inner hose
- ⑨ outer hose
- ⑩ T-adapter
- ⑪ rubber washer
- ⑫ plastic compression fitting
- ⑬ cap

Accessories




Figure 55: Dimensions of Air Cooling System

10.2.5 Right Angle Mirror

The right angle mirror comes in two different versions:

- XXXMIACRAJ right angle mirror as accessory for air purge jacket or air cooling system
- XXXMIACRAJ1 right angle mirror with integrated air purge (not available for LTH sensing heads)




Figure 56: Right Angle Mirror XXXMIACRAJ (left), Right Angle Mirror with Air Purge XXXMIACRAJ1 (right)

The right angle mirror withstands ambient temperatures up to 180°C (356°F). For mounting the right angle mirror (XXXMIACRAJ), see section 10.2.3 [Air Purge Jacket](#), page 61. However, instead of using the front part of the air purge jacket ③, mount the right angle mirror.




Figure 57: Right Angle Mirror (* with Air Purge)

The IR beam length within the right angle mirror is 18 mm (0.7 in.) which needs to be considered for spot size calculations.

10.2.6 Protective Windows

Protective windows can be used to protect the sensing head from dust and other contamination. The protective window can be directly screwed onto the sensing head. It has an outer diameter of 17 mm (0.67 in).

The following table provides an overview of the available windows.

| Order number | Material | Transmission | T _{ambient} |
|--------------|---------------------------------|-----------------------------------|----------------------|
| XXXMIACPW | Zinc Sulfide stainless steel | 0.75 ±0.05 (for LT, G5 models) | 180°C (356°F) |
| XXXMI3ACPWP | Polymer stainless steel | 0.7 ±0.02 (LT models only) | 65°C (149°F) |

Table 5: Available Protective Windows



For correct temperature readings, the transmission of the protective window must be set via the control panel in the communication box. See section 8.2 <Head> Page, page 47!



Figure 58: Protective Window

10.2.7 Close Focus Lens

The close focus lens is designed to get very small measurement spots down to 0.5 mm (0.02 in). The lens should be used for LT models only.

The close focus lens is made from Silicon, with a focal distance of 10 mm (0.39 in) and a transmission factor of 0.75 ± 0.05 (for 8 to 14 μm). It has an outer diameter of 17 mm (0.67 in). The close focus lens can be directly screwed onto the sensing head. It withstands ambient temperatures up to 180°C (356°F).



For correct temperature readings, the transmission of the close focus lens must be set via the control panel in the communication box. See section 8.2 <Head> Page, page 47!

Accessories




Figure 59: Sensing Head with Close Focus Lens (XXXMI3ACCFL)




Figure 60: Spot Size Charts for Close Focus Lens

10.3 Accessories (1M, 2M Heads)

- [Fixed Mounting Bracket \(XXXMI3100FB\)](#)
- [Adjustable Mounting Bracket \(XXXMI3100ADJB\)](#)
- [Air Purge Collar \(XXXMI3100AP\)](#)
- [Right Angle Mirror \(XXXMI3100RAM\)](#)
- [Protective Window \(XXXMI3100PW\)](#)




Figure 61: Overview of available accessories

Accessories

10.3.1 Fixed Mounting Bracket




Figure 62: Dimensions of Fixed Mounting Bracket (XXXMI3100FB)

10.3.2 Adjustable Mounting Bracket




Figure 63: Dimensions of Adjustable Mounting Bracket (XXXMI3100ADJB)

10.3.3 Air Purge Collar

The Air Purge Collar is used to keep dust, moisture, airborne particles, and vapors away from the lens. It can be mounted before or after the bracket. It has the push-in fitting. A 4 mm (0.16 in) outside diameter plastic tubing is recommended to connect the fitting. Air flows into the fitting and out the front aperture. The pressure of air should be 0.6 to 1 bar (8.7 to 15 PSI). Clean, oil free air is recommended.




Figure 64: Dimensions of Air Purge Collar (XXXMI3100AP)

Accessories

10.3.4 Right Angle Mirror

The Right Angle Mirror is used to turn the field of view by 90° against the sensor axis. It is recommended when space limitations or excessive radiation do not allow for direct alignment of the sensor to the target. The mirror must be installed after the bracket and after the Air Purge Collar and screwed in fully. In dusty or contaminated environments, air purging is required to keep the mirror surface clean.




Figure 65: Dimension of Right Angle Mirror (XXXMI3100RAM)



When using the Right Angle Mirror, adjust the emissivity or transmissivity settings downward by 5%. For example, for an object with an emissivity of 0.65, you adjust the value down to 0.62. Or, you can keep the emissivity 0.65 and adjust the transmissivity from 1.0 to 0.95. This correction accounts for energy losses in the mirror.

10.3.5 Protective Window

Protective windows can be used to protect the sensing head from dust and other contamination. The protective window can be directly screwed onto the sensing head.

The following table provides an overview of the available windows.

| Order number | Material | Transmission | T _{ambient} |
|--------------|---------------------------------|-----------------------------------|----------------------|
| XXXMI3100PW | Fused Silica stainless steel | 0.98 ±0.05 (for 1M, 2M models) | 120°C (248°F) |

Table 6: Available Protective Windows



For correct temperature readings, the transmission of the protective window must be set via the control panel in the communication box. See section 8.2 <Head> Page, page 47!




Figure 66: Protective Window

Maintenance

11 Maintenance

Our sales representatives are always at your disposal for questions regarding application assistance, calibration, repair, and solutions to specific problems. Please contact your local sales representative, if you need assistance. In many cases, problems can be solved over the telephone. If you need to return equipment for servicing, calibration, or repair, please call our Service Department for authorization prior to return. Phone numbers are listed at the beginning of this document.

11.1 Troubleshooting Minor Problems

| Symptom | Probable Cause | Solution |
|------------------------|---------------------------|--|
| No output | No power to instrument | Check the power supply |
| Erroneous temperature | Faulty sensor cable | Verify cable continuity |
| Erroneous temperature | Field of view obstruction | Remove the obstruction |
| Erroneous temperature | Window lens | Clean the lens |
| Erroneous temperature | Wrong emissivity | Correct the setting |
| Temperature fluctuates | Wrong signal processing | Correct Peak/Valley Hold or Average settings |
| Temperature fluctuates | No ground for the head | Check wiring / grounding |

Table 7: Troubleshooting

11.2 Fail-Safe Operation

The Fail-Safe system is designed to alert the operator and provide a safe output in case of any system failure. The sensor is designed to shutdown the process in the event of a set-up error, system error, or a failure in the sensor electronics.



The Fail-Safe circuit should never be relied on exclusively to protect critical processes. Other safety devices should also be used to supplement this function!

When an error or failure does occur, the display indicates the possible failure area, and the output circuits automatically adjust to their preset levels. See the following tables:

| Symptom | 0 to 5 V | 0 to 10 V | 0 to 20 mA | 4 to 20 mA |
|--|----------|-----------|-------------|-------------|
| Temperature over range* | 5 V | 10 V | 21 to 24 mA | 21 to 24 mA |
| Temperature under range* | 0 V | 0 V | 0 mA | 2 to 3 mA |
| Head ambient temperature out of range | 5 V | 10 V | 21 to 24 mA | 21 to 24 mA |
| Communication error between head and box | 5 V | 10 V | 21 to 24 mA | 21 to 24 mA |

* related to zoomed temperature range

Table 8: Error Codes for Analog Output

| Symptom | J | K | R | S |
|---------------------------------------|----------------------|----------------------|----------------------|----------------------|
| Temperature over range | > 1200°C (2192°F) | > 1372°C (2502°F) | > 1768°C (3214°F) | > 1768°C (3214°F) |
| Temperature under range | -210°C (-346°F) | -210°C (-346°F) | -50°C (-58°F) | -50°C (-58°F) |
| Head ambient temperature out of range | > 1200°C (2192°F) | > 1372°C (2502°F) | > 1768°C (3214°F) | > 1768°C (3214°F) |

Table 9: Error Codes for Thermocouple Output TC

| Output | Error Code Description |
|--------|--|
| T--- | Communication error between head and box |
| T>>> | Temperature over range |
| T<<< | Temperature under range |

Table 10: Error Codes via Field Bus

| Display | Error Code Description |
|------------------------|---|
| "No sensor" | No sensing head detected |
| "Sensing head #n lost" | Communication error between head and box |
| ">" | Temperature over top range* e.g. ">600°C" |
| "<" | Temperature under bottom range* e.g. "<-40°C" |

* related to full measurement range

Table 11: Error Codes for LCD Display

11.3 Cleaning the Lens

Keep the lens clean at all times. Care should be taken when cleaning the lens. To clean the window, do the following:

1. Lightly blow off loose particles with "canned" air (used for cleaning computer equipment) or a small squeeze bellows (used for cleaning camera lenses).
2. Gently brush off any remaining particles with a soft camel hair brush or a soft lens tissue (available from camera supply stores).
3. Clean remaining "dirt" using a cotton swab or soft lens tissue dampened in distilled water. Do not scratch the surface.

For finger prints or other grease, use any of the following:

- Denatured alcohol
- Ethanol
- Kodak lens cleaner

Apply one of the above to the lens. Wipe gently with a soft, clean cloth until you see colors on the surface, then allow to air dry. Do not wipe the surface dry, this may scratch the surface.

If silicones (used in hand creams) get on the window, gently wipe the surface with Hexane. Allow to air dry.


Maintenance



Do not use any ammonia or any cleaners containing ammonia to clean the lens. This may result in permanent damage to the lens' surface!

11.4 Sensing Head Exchange

To exchange a sensing head, the following procedure is required:

1. Disconnect power to the box.
2. Disconnect all head wires from the box terminal.
3. Power the box.
4. The alarm indicator of the box starts to blink indicating a lost sensing head.
5. Press the  button to navigate to the head page indicating a lost sensing head.
6. Decide whether you want to select:
 - a) <Remove Yes>: to **remove** the head permanently from the box by loosing all head parameters (head address de-allocated for other heads, alarm condition is reset) – the next head connected later will be detected as a new head and automatically assigned to a free head address.
 - or
 - b) <Remove No>: to **keep** the head assigned to the box by saving all head parameters for a future use without the need to parameterize that head again (head address reserved for that individual head, alarm condition is kept) – the same head connected later will be detected as a known head by keeping the previous address.

12 DataTemp Multidrop Software

12.1 Software Features

DataTemp Multidrop DTMD provides sensor setup, remote monitoring, and simple data logging for analysis or to meet quality record-keeping requirements. Additional features configurable with DTMD Software:

- Eight-position “recipe” table that can be easily interfaced to an external control system
- External reset signal input FTC for signal processing
- External inputs FTC for analog emissivity adjustment or background radiation compensation
- Remote digital communication and control of up to 32 sensors in an RS485 multidrop configuration

For more detailed information, see the comprehensive help feature in the DTMD software.

12.2 PC Requirements

- PC with Windows 2000/XP/Vista/Win7, 64 MB RAM memory
- about 10 Mb free memory on the hard disc for program files
- USB port with recommended USB/RS485 adapter (available as accessory), see section 10.1.2 [USB/RS485 Adapter](#), page 57.

12.3 USB Driver Installation

Before running the DTMD Software the installation of an adequate USB driver is required, see section 5.6 [USB](#), page 34.

12.4 Software Launch

Make sure any sensor is turned on and the USB driver is installed before running DTMD software. The Startup Wizard runs the first time you use the program. Please note:

- The wizard shows active COM ports only!
- The sensor requires the selection of <ASCII protocol>!
- The DTMD software communicates to Comm Boxes only! A multidrop installation is related to a network with multiple Comm Boxes and not to a multiple head system with one Comm box only!

RS485

13 RS485

The RS485 serial interface is used for long distances up to 1200 m (4000 ft) or for networked communication boxes.


To connect the RS485 interface to a standard computer you should use a dedicated adapter, see section 10.1.2 [USB/RS485 Adapter](#), page 57. The RS485 interface allows the communication either via the standard Multidrop Software or directly via dedicated ASCII commands, see section 18 [ASCII Programming](#), page 103.

Specification:

- Physical layer: RS485, 2 wire, half-duplex, electrically isolated
- Baud rate: 9.6, 19.2, 38.4, 57.6, 115.2 kBit/s
- Settings: 8 data bits, 1 stop bit, no parity, flow control: none (half duplex mode)
- Connection: terminal
- Address range: 1 to 32
0 for stand-alone unit or broadcast transmission

13.1 Wiring


13.1.1 Comm Box (metal)



| Terminal | RS485 |
|----------|---------------------------------|
| A | positive signal (or RxA, or D+) |
| B | negative signal (or RxB, or D-) |

Figure 67: RS485 Terminal for Comm Box (metal)

13.1.2 Comm Box (DIN)



| Pin | RS485 |
|---------------------|---|
| 1 | n.a. |
| 2 | Shield |
| 3 (negative signal) | A2 (not supported while termination "on") |
| 4 (positive signal) | B2 (not supported while termination "on") |
| 5 (negative signal) | A1 |
| 6 (positive signal) | B1 |
| 7 | Shield |
| 8 | n.a. |
| 9 | n.a. |
| LED, yellow | n.a. |
| LED, red | n.a. |

Figure 68: Terminal for Comm Box (DIN 6TE)

13.2 ASCII Programming

For the programming details, see section 18 [ASCII Programming](#), page 103.

Profibus

14 Profibus

Profibus DP-V0 defines a cyclical data exchange between a master (e.g. a PLC) and a slave (MI3 sensor). At start-up first an array of parameters (Profibus specific data) is sent from the master to the slave, followed by an array with the configuration (sensor specific presetting's taken from the GSD file) also sent from the master to the slave.

After start-up the bus switches to the data exchange state. In this state in- and output data gets exchanged cyclically between master and slave. The **input data** is sent from the slave to the master and contains mainly the measured temperatures of the MI3 sensor, see section 14.2.2 [Input Data](#), page 84. The **output data** is sent from the master to the slave and contains a set of selected sensor parameters, see section 14.2.3 [Output Data](#), page 85. In case of an error in start-up phase or during data exchange **diagnostic data** is sent to the master, see section 14.2.4 [Diagnose Data](#), page 85.


Each Profibus device comes with a device description file (GSD file) which is read by the programming software of the master to define the slave.

Specification:

| | |
|-----------------|--|
| Version: | Profibus DP-V0 |
| Physical layer: | RS485, 2 wire, electrically isolated |
| Baud rate: | 9.6 kBit/s to 12 MBit/s (automatic negotiated) |
| Connection | terminal or Sub-D or M12 |
| Address range: | 1 to 125 (for the Profibus device) |
| ID | 0D36 |
| GSD file | "RAY_0D36.gsd" |
| Head support | up to 8 sensing heads (MI3 or MI3100) |

14.1 Wiring

14.1.1 Comm Box (metal)



| X1 Pin Terminal | Profibus |
|-----------------|--|
| 1 | A (negative signal) |
| 2 | B (positive signal) |
| 3 | Shield |
| 4 | GND (output, used for external termination) |
| 5 | n.a. |
| 6 | + 5 V (output, used for external termination) |
| LED, yellow | Blinking: 0.5 Hz: parameters / 1 Hz: configuration |
| LED, red | ON: data-exchange |

Figure 69: Profibus Terminal for Comm Box (metal)

A Sub-D female connector or a M12 female connector can be ordered separately for Profibus. The M12 connector is B-coded. Please note the Sub-D connector is not IP rated!




Figure 70: Sub-D Connector (...P2) and M12 Connector (...P1)

| Profibus | Sub-D Pin (cable color) | M12 Pin (cable color) |
|---------------------|----------------------------|--------------------------|
| A (negative signal) | 8 (green) | 2 (green) |
| B (positive signal) | 3 (yellow) | 4 (red) |
| Shield | | |
| GND | 5 (white) | 3 (blue) |
| n.a. | | |
| + 5 V | 6 (brown) | 1 (brown) |

Figure 71: Profibus Pin Assignment for Sub-D / M12 Connector



The termination for Profibus networks must be realized externally by the user!





Figure 72: Exemplary Network with External Termination

Profibus

14.1.2 Comm Box (DIN)



| Pin | Profibus |
|---------------------|--|
| 1 | n.a. |
| 2 | Shield |
| 3 (negative signal) | A2 (not supported while termination "on") |
| 4 (positive signal) | B2 (not supported while termination "on") |
| 5 (negative signal) | A1 |
| 6 (positive signal) | B1 |
| 7 | Shield |
| 8 | n.a. |
| 9 | DE |
| LED, yellow | Blinking: 0.5 Hz: parameters / 1 Hz: configuration |
| LED, red | ON: data-exchange |

Figure 73: Terminal for Comm Box (DIN 6TE)

14.2 Programming

14.2.1 Parameter Data

| Byte | Address without offset | Description | Format | Range |
|---------|------------------------|-----------------------------------|--|-------------------------------|
| 0 to 6 | | Fix | | |
| 7 | | DP-V1 Status1 | | |
| 8 | | DP-V1 Status2 | | |
| 9 | | DP-V1 Status3 | | |
| 10 | 3 | Temp. unit | 67=°C, 70=°F | 67 or 70 |
| 11 | 4 | Reserved | | |
| 12, 13 | 5 | Bottom temp. of output 1 | in 0.1°C /°F | |
| 14, 15 | 7 | Top temp. of output 1 | in 0.1°C /°F | |
| 16, 17 | 9 | Bottom temp. of output 2 | in 0.1°C /°F | |
| 18, 19 | 11 | Top temp. of output 2 | in 0.1°C /°F | |
| 20 | 13 | Source (head) for output-1 | head number | 1...8 |
| 21 | 14 | Type of source for output-1 | 73 = I (internal temp.) 84 = T (object temp.) | 73 or 84 |
| 22 | 15 | Analog output mode 1 | 5 = TCJ 6 = TCK 7 = TCR 8 = TCS 9 = 0...5V 10 = 0...10V 99 = tristate (disabled) | 5...10, 99 default: 9 |
| 23 | 16 | Source (head) for output-2 | head number | 1...8 |
| 24 | 17 | Type of source for output-2 | 73 = I (internal temp.), 84 = T (object temp.) | 73 or 84 |
| 25 | 18 | Analog output mode 2 | 0 = 0...20mA 4 = 4...20mA 9 = 0...5V 10 = 0...10V 99 = tristate (disabled) | 0, 4, 9, 10, 99 default: 4 |
| 26...39 | | reserved | | |
| 40...43 | | reserved | | |
| 44, 45 | 37 | Emissivity Head_1 | * 1000 (0.9 → 900) | 100 ... 1100 |
| 46, 47 | 39 | Transmissivity Head_1 | * 1000 (1.0 → 1000) | 100 ... 1000 |
| 48, 49 | 41 | Averaging time Head_1 | * 0.1s (1s → 10) | 0 ...9990 |
| 50, 51 | 43 | Peak hold time Head_1 | * 0.1s (1s → 10) | 0 ...9990 |
| 52, 53 | 45 | Valley hold time Head_1 | * 0.1s (1s → 10) | 0 ...9990 |
| 54, 55 | 47 | Ambient temp. Head_1 | in °C /°F | dev. range min.. max |
| 56, 57 | 49 | Setpoint relay Head_1 | in °C /°F | dev. range min.. max |
| 58 | 51 | Relay alarm output control Head_1 | 0 = off 1 = target temp. 2 = internal temp. | 0, 1 or 2 |

Profibus

| Byte | Address without offset | Description | Format | Range |
|---------|------------------------|------------------------------------|----------------------------------|--------|
| 59 | 52 | Laser Head_1 | 0 = off, 1 = on, 2 = flashing | 0 or 1 |
| 60...63 | | reserved, for future consideration | | |
| 64 ... | | Head_2 | | |
| 84 ... | | Head_3 | | |
| 104 ... | | Head_4 | | |
| 124 ... | | Head_5 | | |
| 144 ... | | Head_6 | | |
| 164 ... | | Head_7 | | |
| 184 ... | | Head_8 | | |

14.2.2 Input Data

The input data consists of modules that have a fixed position in the data field. There are two types of modules: <Box data> and <Data for one head>.

- Module <**Box data**> consists of one byte in which bit0 gives the trigger state (configuration 0x12).
- Module <**Head data**> consists of two bytes object temperature and two bytes head temperature (configuration 0x51). The format is 1/10 °C/°F.

| Address | Description |
|---------|--------------------------------|
| 0 | Box data: Trigger state |
| 1, 2 | Box data: Internal temperature |
| 3, 4 | Head 1: Object temperature |
| 5, 6 | Head 1: Head temperature |
| 7, 8 | Head 2: Object temperature |
| 9, 10 | Head 2: Head temperature |
| 11, 12 | Head 3: Object temperature |
| 13, 14 | Head 3: Head temperature |
| 15, 16 | Head 4: Object temperature |
| 17, 18 | Head 4: Head temperature |
| 19, 20 | Head 5: Object temperature |
| 21, 22 | Head 5: Head temperature |
| 23, 24 | Head 6: Object temperature |
| 25, 26 | Head 6: Head temperature |
| 27, 28 | Head 7: Object temperature |
| 29, 30 | Head 7: Head temperature |
| 31, 32 | Head 8: Object temperature |
| 33, 34 | Head 8: Head temperature |

The slave expects one <Box data> module configured at the first position, followed by <Head data> modules. Any other configuration will cause a configuration error.

The input data length gets calculated from the count of configured modules. So, if only one head is connected and configured then only seven bytes are transferred. If all heads (maximal eight) are connected and configured then 35 bytes are transferred. If only one head is connected but eight heads are configured then also 35 bytes are transferred.

14.2.3 Output Data

The device does not have output data in the original meaning. But the output data may be used to change the initialization of the device (which was set once at start-up) when the bus is in data exchange mode.

To do so the following structure is defined:

| Address | Description |
|---------|----------------------|
| 0 | Type of parameter |
| 1, 2 | Parameter for Head 1 |
| 3, 4 | Parameter for Head 2 |
| 5, 6 | Parameter for Head 3 |
| 7, 8 | Parameter for Head 4 |
| 9, 10 | Parameter for Head 5 |
| 11, 12 | Parameter for Head 6 |
| 13, 14 | Parameter for Head 7 |
| 15, 16 | Parameter for Head 8 |

The <Type of parameter> comes with the format described in section 14.2.1 [Parameter Data](#), page 83, and can be set to the following parameters:

| Number of type | Description |
|----------------|--|
| 0 | do not change anything |
| 1 | emissivity |
| 2 | laser |
| 3 | ambient temperature (ambient temperature compensation) |
| 4 | averaging time |
| 5 | peak hold time |
| 6 | valley hold time |
| 7 | set point for the relay |

If <Type of parameter> is set to 0 then the output data gets ignored. So it should be set to 0 as default.

Attention: You should be aware that always all heads are updated! So you have to set all eight (or as much as heads are connected) parameters to the correct value!

14.2.4 Diagnose Data

The device uses the first 32 bytes of the Identifier Related Diagnosis.

The first 6 bytes consist of Standard Diagnosis dedicated to bus parameters. In this field byte 4 and 5 give the unit identifier (0D36 in our case).

Profibus

| Byte | Description |
|-------------------------|---|
| 6 | size of the diagnosis |
| 7...9 | reserved |
| 10 (0x0A) | highest address of the connected heads → up to which index the user data is valid |
| 11 (0x0B) | box error code |
| 12 ... 22 (0x0C ... 16) | last MI3-command which created an error as answer; ASCII code |
| 23 (0x17) | head_1 error code |
| 24 (0x18) | head_2 error code |
| 25 (0x19) | head_3 error code |
| 26 (0x1A) | head_4 error code |
| 27 (0x1B) | head_5 error code |
| 28 (0x1C) | head_6 error code |
| 29 (0x1D) | head_7 error code |
| 30 (0x1E) | head_8 error code |

Table 12: Diagnose Data

| Bit | Description |
|-----|--------------------------------------|
| 0 | Self-test error |
| 1 | Box ambient temperature out of range |
| 2 | Sensing head communication error |
| 3 | Parameter error |
| 4 | Register write error |

Table 13: Error Bits of Box Diagnose

| Bit | Description |
|-----|--|
| 0 | Object temperature out of range |
| 1 | Ambient (internal) temperature out of range |
| 2 | Parameter error |
| 3 | Register write error |
| 4 | Self-test error |
| 5 | reserved |
| 6 | Head registered but not connected → cable break? |

Table 14: Error Bits of Head Diagnose




Figure 74: Diagnose Data without Errors




Figure 75: Diagnose Data with Error “Cable Break at Head 1”

Modbus

15 Modbus

The Modbus protocol follows the master/slave model. One master controls one or more slaves. Typically, the master sends a request to a slave, which in turn sends a response. The request/response mechanism is called a transaction. Requests and responses are also referred to as messages.


Specification:

| | |
|-----------------|--------------------------------------|
| Version: | Modbus serial line (RS485) |
| Mode: | RTU (Remote Terminal Unit) |
| Physical layer: | RS485, 2 wire, electrically isolated |
| Baud rate: | 9.6, 19.2, 38.4, 57.6, 115.2 kBit/s |
| Connection | terminal |
| Address range: | 1 to 247 (for the Modbus device) |
| Parity | even |

The detailed Modbus specification can be found under <http://www.modbus.org/>.

15.1 Wiring

15.1.1 Comm Box (metal)




| X1 Pin Terminal | Modbus |
|-----------------|---|
| 1 | D0 (negative signal) |
| 2 | D1 (positive signal) |
| 3 | Shield |
| 4 | GND (output, used for external termination) |
| 5 | n.a. |
| 6 | + 5 V (output, used for external termination) |
| LED, yellow | ON while communicating |
| LED, red | Error |

Figure 76: Modbus Terminal for Comm Box (metal)



The termination for Modbus networks must be realized externally by the user!

15.1.2 Comm Box (DIN)



| Pin | Modbus |
|---------------------|---|
| 1 | GND (output) |
| 2 | Shield |
| 3 (negative signal) | D0_2 (not supported while termination "on") |
| 4 (positive signal) | D1_2 (not supported while termination "on") |
| 5 (negative signal) | D0_1 |
| 6 (positive signal) | D1_1 |
| 7 | Shield |
| 8 | GND (output) |
| 9 | n.a. |
| LED, yellow | Communication |
| LED, red | Error |

Figure 77: Terminal for Comm Box (DIN 6TE)

Modbus

15.2 Programming



Functions codes and start addresses are listed in decimal.

15.2.1 Supported Functions

| Function code | Modbus Function | Description |
|---------------|--------------------------|----------------------|
| 01 | Read Coils | Read n bits |
| 02 | Read Discrete Inputs | Read n bits |
| 03 | Read Holding Registers | Read n 16 bit words |
| 04 | Read Input Registers | Read n 16 bit words |
| 06 | Write Single Register | Write a 16 bit word |
| 16 | Write Multiple Registers | Write n 16 bit words |

15.2.2 Parameter Data

32 bit registers are transmitted in full Big-Endian mode, meaning most significant word is transmitted first, least significant word is transmitted last. The byte order within a word is Big-Endian as well.

While some registers hold integer values, there are some holding floating points. The interchange formats of the IEEE-754 standard for Floating-Point Arithmetic is used for representing floating points.

The following table lists all parameters, its content, their formats and usage.

15.2.2.1 Box Parameter

<k> ... number of output channel, depending on the number of physically installed output channels of the Comm Box.

Items (registers, discretives or coils) are addressed starting at zero. Therefore items numbered 1-10000 are addressed as 0-9999.

| Start address | Size [bits] | Modbus Access | Data Type | Content | Values | MI3[M] command |
|---------------|-------------|------------------|-----------|-----------------------------|--|----------------|
| 1 | 16 | input register | short | error code for last request | 0: no error 1: value out of range 2: illegal head number 3: illegal analog output number 4: illegal output mode 5: output disabled error 99: unspecified error | - |
| 10 | 64 | input register | hex | Serial number | e.g. 98123 | XV |
| 20 | 64 | input register | string | Unit identification | e.g. MI3COMM | XU |
| 30 | 64 | input register | string | Box Firmware Revision | e.g. 2.10 | XR |
| 40 | 16 | input register | short | Modbus slave address | 1 .. 247 | XA |
| 50 | 32 | input register | string | Box special | e.g. RAY, LAS | DS |
| 60 | 32 | holding register | integer | Baud rate RS485 | 9600, 19200, 38400, 57600, 115200 | BR |
| 70 | 16 | holding register | char | Temperature Unit | 0x43 ('C'), 0x46 ('F') | U |
| 80 | 32 | input register | float | Box Temperature | | XJ |
| 90 | 16 | holding register | short | Switch panel lock | 0: unlocked, 1: locked | J |

Modbus

| Start address | Size [bits] | Modbus Access | Data Type | Content | Values | MI3[M] command |
|---------------|-------------|------------------|-----------|------------------------------------|--|----------------|
| 100 | 8 | discretes input | bit field | Get connected heads | bit 0: head 1 .. bit 7: head 8 bit high: head connected bit low: head disconnected | HC |
| 110 | 8 | discretes input | bit field | Get registered heads | bit 0: head 1 .. bit 7: head 8 bit high: head registered bit low: head not registered | HCR |
| 120 | 16 | holding register | short | Laser control (only with laser) | 0: off, 1: on | XL |
| 130 | 16 | holding register | short | Relay alarm output control | 0: off 1: on 2: norm. open 3: norm. closed | KB |
| 410 | 32 | input register | float | analog input 1 value | 0 .. 5 V | TV1I |
| 420 | 32 | input register | float | analog input 2 value | 0 .. 5 V | TV2I |
| 430 | 16 | input register | short | Trigger | 0: off 1: on | XT |
| 5<k>0 | 16 | holding register | short | analog output k mode | 5: TCJ 6: TCK 7: TCR 8: TCS 9: 0...5 V 10: 0...10 V 99: disable (tristate) | XO<k>0 |
| 5<k>1 | 16 | holding register | short | analog output k source | head number or 0: fixed value from 5<k>3 | O<k>0 |
| 5<k>2 | 16 | holding register | short | analog output k source parameter | 1: internal temp. of 5<k>1 2: object temp. of 5<k>1 | O<k>0 |
| 5<k>3 | 32 | holding register | float | analog output k fixed temp. value | value within range set in address 5<k>0 | - |
| 5<k>5 | 32 | holding register | float | analog output k bottom temp. value | device bottom temp. .. device top temp. | L<k>0 |
| 5<k>7 | 32 | holding register | float | analog output k top temp. value | device bottom temp. .. device top temp. | H<k>0 |

Modbus

15.2.2.2 Head Parameter

<n> ... head number, depending on the registered heads

| Starting address | Size [bits] | Modbus Access | Data Type | Content | Values | MI3[M] command |
|------------------|-------------|------------------|-----------|---|---|----------------|
| <n>005 | 8 | discretes input | bit field | Head Status | bit0: Temperature Unit (0: deg. C, 1: deg. F) bit1: Object temperature out of range bit2: Ambient temperature out of range bit3: Parameter error bit4: Register write error bit5: Self-test error bit7: Background temp. compensation (0: off, 1: on) | HEC |
| <n>010 | 64 | input register | hex | Head serial number | e.g. 10C02752 | HN |
| <n>020 | 64 | input register | string | Head identification | e.g. "MI310LTS" | HI |
| <n>030 | 64 | input register | string | Firmware Rev. Head | e.g. 2.10 | HV |
| <n>040 | 16 | input register | short | Head Address | 1 .. 9 | HA |
| <n>050 | 32 | input register | string | Head special | e.g. !99HSRAY (read only) | HS |
| <n>060 | 32 | input register | float | Device Bottom range | -40 .. 1300 [°C] | XB |
| <n>070 | 32 | input register | float | Device High range | -40 .. 1300 [°C] | XH |
| <n>080 | 32 | input register | float | Target temperature | number within range <n>060 .. <n>070 | T |
| <n>090 | 32 | input register | float | internal sensor temp. | | I |
| <n>100 | 32 | holding register | float | Ambient background temp. compensation | | A |
| <n>110 | 32 | holding register | float | Advanced hold with average | 0 .. 999.9 [s] | AA |
| <n>120 | 16 | holding register | short | Control ambient background temp. compensation | 0: sensor temp. 1: via number 2: via ext. input | AC |
| <n>130 | 16 | holding register | short | Laser control (only with laser device) | 0: OFF, 1: ON | HL |
| <n>140 | 16 | holding register | short | Relay alarm output control | 0: off 1: Target 2: Ambient | KH |
| <n>150 | 32 | holding register | float | Advanced hold threshold | number within range <n>060 .. <n>070 | C |
| <n>160 | 32 | input register | float | Currently calculated emissivity | 0.1 .. 1.1 | CE |
| <n>170 | 32 | input register | float | Current calculation setpoint / relay function | number within range <n>060 .. <n>070 | CS |
| <n>180 | 32 | holding register | float | Sensor Gain | 0.8 .. 1.2 | DG |
| <n>190 | 32 | holding register | float | Sensor Offset | -200 C .. +200 C | DO |
| <n>200 | 32 | holding register | float | Emissivity internal | 0.1 .. 1.1 | E |
| <n>210 | 16 | holding register | short | Presel. Emissivity Pointer | 0 .. 7 | EP |

Modbus

| Starting address | Size [bits] | Modbus Access | Data Type | Content | Values | MI3[M] command |
|------------------|-------------|------------------|-----------|-------------------------|---|----------------|
| <n>220 | 16 | holding register | char | Emissivity Source | I: internal command E: ext. input (0V .. 5V) D: digital selected FTC1-3 | ES |
| <n>230 | 32 | holding register | float | Presel. Emissivity | 0.1 .. 1.1 | EV |
| <n>240 | 32 | holding register | float | Valley hold time | 0.0 .. 998.9s (999: infinite) | F |
| <n>250 | 32 | holding register | float | Average time | 0.0 .. 999.0 [seconds] | G |
| <n>260 | 32 | holding register | float | Peak hold time | 0.0 .. 998.9 [seconds] (999: infinite) | P |
| <n>270 | 32 | input register | integer | Power / AD value | | Q |
| <n>280 | 32 | holding register | float | Presel. Setpoint | | SV |
| <n>290 | 32 | holding register | float | Transmissivity | 0.1 .. 1.0 | XG |
| <n>300 | 16 | holding register | short | FTC3 trigger/hold | 1: trigger, 2: hold | XN |
| <n>310 | 32 | holding register | float | Setpoint relay function | number within range <n>060 .. <n>070 | XS |
| <n>320 | 32 | holding register | float | Adv. hold hysteresis | | XY |

Ethernet

16 Ethernet

Specification:

Ethernet: 10/100 MBit/s, Auto-Negotiation
Protocol: TCP/IP version 4
http server capability supporting up to 8 sensing heads
Connection: M12 or RJ45
electrically isolated




An opened Ethernet port will be closed automatically by the MI3 communication box if there is no data traffic within 120 s!

16.1 Wiring

16.1.1 Comm Box (metal)

The connector on the box side is an M12 plug-in connector, 4 pin D-coded, suited for industrial Ethernet with IP67 protection rate with a screw retention feature. An M12/RJ45 connector adapter cable is available at a length of 7.5 m / 25 ft withstanding up to 70°C/158°F for the ambient temperature (XXXETHLTCB).




| M12-Pin | |
|---------|-----|
| 1 | TX+ |
| 2 | RX+ |
| 3 | TX- |
| 4 | RX- |

Figure 78: M12 Connector Socket and Pin Assignment

16.1.2 Comm Box (DIN)

The connector on the box side is a standardized RJ45 connector.



| RJ45-Pin | |
|----------|-----|
| 1 | TX+ |
| 3 | RX+ |
| 2 | TX- |
| 6 | RX- |

Figure 79: RJ45 Connector Socket and Pin Assignment

16.2 ASCII Programming

For the programming details, see section 18 [ASCII Programming](#), page 103.

16.3 http Server

The MI3 Comm Box with Ethernet provides a built-in http server for one or more client computers based on the http protocol within an Intranet. For getting/setting the network address through the control panel of the Comm Box, see section 8.3 <Box Setup> Page, page 48.

The web site landing page for a client is to be seen in the next figure.

Raytek
A Fluke Company

Noncontact Infrared Temperature Measurement

TEMPERATURE SENSOR WEB MONITOR

Box Model: MI3MCOMME
S/N: 12345678
FW Revision: 2.19
Internal temperature, °C : 20.1

Sensing head: 1
Head Model: MI310LTH
S/N: 12706680
FW Revision: 2.06

| Sensing head(s): | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 |
|----------------------------|------|------|------|------|------|------|------|------|
| Object temperature, °C : | 21.2 | 22.5 | 22.2 | 21.5 | 21.3 | 21.8 | 21.6 | 22.2 |
| Internal temperature, °C : | 22.3 | 22.6 | 22.7 | 22.5 | 22.3 | 22.4 | 22.0 | 22.8 |
| Status: | ok. | ok. | ok. | ok. | ok. | ok. | ok. | ok. |

DATA LOGGING stopped
Interval, (s): 1 Set
Start Stop
GET DATA
Password: ●●●●
New password: Set

- Box Datasheet (EN)
- Box Datenblatt (DE)
- Boitiers Fiche technique (FR)
- Head Datasheet (EN)
- Kopf Datenblatt (DE)
- Capteurs Fiche technique (FR)
- MI3 QuickStart
- MI3M QuickStart
- MI3 Manual (EN)

Raytek website | Support
© Raytek Corporation 2012 - 2013

Figure 80: Landing Page for http Clients

Beside some basic information for the MI3 communication box (box model, serial number S/N, firmware revision, and current internal temperature), all sensing heads connected to the box are listed on that page. The head number is mouse sensitive and provides additional information to the head (head model, serial number S/N, firmware revision).

Each sensing head provides the current measured object temperature, the internal temperature of the head itself, and the operating status differentiated by the following:

- <ok.> operating status ok
- <error> communication error detected (e.g. in case of a cable break or a power loss)
- <alarm> alarm condition fulfilled (current object temperature versus the alarm threshold)

The one-click access area on the right side of the page provides supporting documents which are stored on the box itself.

Ethernet

16.3.1 Data Logging

With the data logging function on the MI3 communication box you can record the temperature values over time directly in the internal memory of the box.

Pressing on the <Start> button triggers the data storage in accordance to the defined time interval, given in seconds. The internal memory of the box is 64 MB in size allowing a permanent data storage over 24 days at a saving interval of 1 s independent from the number of connected heads.



Pressing on the <Start> button again will initiate a new recording session by overwriting the previously stored data.

Clicking on the <Get data> button opens a dialog box to download the previously stored data as *.dat file. The data logging file comes in a standard ASCII text format accessible e.g. by means of the Windows Notepad.




Figure 81: Example for a logging file with recorded data for the measured object temperature and the internal temperature for sensing head 1

The data logging function is password protectable to ensure data consistency while having multiple clients connected to the MI3 http server.



A valid password accepts digits only! The factory presetting is 1234.

17 Profinet

The Profinet fieldbus MI3 communication box (called fieldbus communicator) maps the object and internal temperature of all pyrometer modules on Profinet IO. In the initialization phase, the fieldbus communicator determines the physical structure of the node and creates a local process image with all pyrometers.

The diagnostics concept is based on channel specific diagnostic messages that are mapped to the respective alarms. Coding is done according to IEC 61158 Profinet IO.

Specification:

| | |
|--------------------|---|
| Type: | Profibus IO |
| Conformance class: | A |
| Real-Time class: | 1 (RT) and the Real-Time class UDP |
| Connection: | M12 or RJ45 electrically isolated |
| Transfer speed | 100 Mbit/s full-duplex, also with auto-negotiation |
| Head support | up to 8 pyrometer modules (MI3 or MI3100 sensing heads) Configurable substitute value behavior for each pyrometer module in the event of failure |
| Power supply: | 12 to 24 V (500 mA max.) |

17.1 Wiring

Details for the wiring you can find under Ethernet in section 16.1 [Wiring](#), page 94.

17.1.1 Status LED

The status LED's are relevant for the Comm Box (DIN) only signaling the following operating states.

| yellow LED | green LED | Description |
|------------|-----------|--|
| off | off | No communication |
| on | 1 Hz | Waiting for configuration and parameters |
| on | on | Device is in data-exchange mode |

Figure 82: Status LED's for the Comm Box (DIN)

Profinet

17.2 Programming

17.2.1 I/O Device Configuration

The fieldbus communicator takes over the task of the I/O device in Profinet IO. Selecting the Pyrometer modules for the process data exchange and defining the time grid is done when configuring the I/O controller. The configuration and parameter setting of the fieldbus communicator and Pyrometer modules is based on the device's GSD file.

17.2.1.1 GSD File

Under Profinet IO, the features of the devices are described by the manufacturer in form of a GSD file and made available to the user. The GSD file for the MI3 IO device:

GSDML-V2.25-Raytek-MI3-20130221

17.2.1.2 Configuration

The I/O device is configured in accordance with the physical arrangement of the node (slot oriented). Module slot 0 contains the fieldbus communicator in its function as station substitute. It does not deliver process data itself (except its own internal temperature and trigger input state), but provides the parameters required to perform overall setting of the I/O device.

Slot 1 to max. 8 reflect the physical arrangement of the pyrometer modules that deliver a part of the process and diagnostics data. All specific information on the relevant module is contained in the associated GSD file.

17.2.2 Parameter Setting

Setting the parameters of both the fieldbus communicator and the connected pyrometer module is performed via "record data" sets. All modules allow the diagnostics message to be locked or released. Once all settings have been made, the I/O device signals that it is ready to send cyclic productive data.

17.2.2.1 Station Parameters

The parameters of the station substitute are used to set the overall settings of the Profinet IO node. Some of the setting are used in the modules as default settings and can be optionally overwritten within the module configuration.

| Parameter | Description | Setting |
|--------------------------------|---|--------------------------------|
| Temperature unit | Set the temperature unit | Celsius |
| | | Fahrenheit |
| Message diagnostics | The diagnostics information of all modules are not transferred to the Profinet I/O controller | message inactive |
| | transferred to the Profinet I/O controller | diagnostic alarm |
| Value behavior on module fault | All process data are immediately reset | set process data to zero |
| | All process data maintain the last status | set process data to last value |
| Bottom temperature of output 1 | Set Bottom temperature of analog output 1 | 0...3000°C /°F |
| Top temperature of output 1 | Set Top temperature of analog output 1 | 0...3000°C /°F |
| Bottom temperature of output 2 | Set Bottom temperature of analog output 2 | 0...3000°C /°F |
| Top temperature of output 2 | Set Top temperature of analog output 2 | 0...3000°C /°F |
| Source (head) for output 1 | Set head number | 1..8 |
| Source (type) for output 1 | Set type of source for output 1 | Internal temperature |
| | | Object temperature |
| Analog output 1 mode | Set output mode | TCJ |
| | | TCK |
| | | TCR |
| | | TCS |
| | | 0...5V |
| | | 0...10V |
| | | tristate (disabled) |
| Source (head) for output 2 | Set head number | 1..8 |
| Source (type) for output 2 | Set type of source for output 2 | Internal temperature |
| | | Object temperature |
| Analog output 2 mode | Set output mode | 0...20mA |
| | | 4...20mA |
| | | 0...5V |
| | | 0...10V |
| | | tristate (disabled) |

Profinet

17.2.2.2 Pyrometer Module Parameters

Certain characteristics of some pyrometer modules can be parameterized during the configuration.

| Parameter | Description | Setting |
|----------------------------|--|---|
| Pyrometer number | Set the number of pyrometer | 1..8 |
| Emissivity | * 1000 (0.9 → 900) | 100 ... 1100 |
| Transmissivity | * 1000 (1.0 → 1000) | 100 ... 1000 |
| Averaging time | * 0.1s (1s → 10) | 0 ...9990 |
| Peak hold time | * 0.1s (1s → 10) | 0 ...9990 |
| Valley hold time | * 0.1s (1s → 10) | 0 ...9990 |
| Ambient compensation | Turn ambient compensation | on / off |
| Ambient temperature | in °C /°F | dev. range min.. max |
| Setpoint relay | in °C /°F | dev. range min.. max |
| Relay alarm output control | | off, target temp., internal temp. |
| Laser control | | off / on / flashing |
| Message diagnostics alarm | The diagnostics information of pyrometer is not transferred to the Profinet I/O controller | message inactive |
| | is transferred to the Profinet I/O controller | message active |
| Message process alarm | The process alarm of pyrometer is not transferred to the Profinet I/O controller | message inactive |
| | is transferred to the Profinet I/O controller | message active |
| Message return of module | The return of pyrometer is not transferred to the Profinet I/O controller | message inactive |
| | is transferred to the Profinet I/O controller | message active |

17.2.3 Input Data Structure

17.2.3.1 Station Input data

The input data length is 5 Byte.

| Address without offset | Length | Format | Value |
|------------------------|--------|-----------------------------|-------------------------------------|
| 0 | 1 Byte | Bit0 | Trigger state (0 – reset, 1 – set) |
| 1 | 4 Byte | Real (Big Endian, Motorola) | Internal temperature of the station |

17.2.3.2 Pyrometer Module Input Data

The input data length of pyrometer module is 8 Byte.

| Address without offset | Length | Format | Value |
|------------------------|--------|-----------------------------|----------------------|
| 0 | 4 Byte | Real (Big Endian, Motorola) | Object temperature |
| 4 | 4 Byte | Real (Big Endian, Motorola) | Internal temperature |

17.2.4 Output Data Structure

The pyrometer module provides output data. The output data length of pyrometer module is 3 Byte. The output data may be used to change the initialization of the device (which was set once at start-up) when the bus is in data exchange mode.

To do so the following structure is defined:

| Address without offset | Length | Format | Value |
|------------------------|--------|--------------------------------|-------------------|
| 0 | 1 Byte | Byte (Big Endian, Motorola) | Type of parameter |
| 1, 2 | 2 Byte | Integer (Big Endian, Motorola) | Parameter |

The <Type of parameter> gives the meaning of the following parameters (with the same format as described in section 17.2.2.2 [Pyrometer Module Parameters](#), page 100):

| Number of type | Meaning |
|----------------|--|
| 0 | ignore output data |
| 1 | emissivity |
| 2 | laser |
| 3 | ambient temperature (for ambient temperature compensation) |
| 4 | averaging time |
| 5 | peak hold time |
| 6 | valley hold time |
| 7 | set point for the relay |

If <Type of parameter> is set to 0 then the output data gets ignored. So it should be set to 0 as default.

17.2.5 Diagnostics

The diagnostics information of the fieldbus communicator can be read out acyclically using standard diagnostics data sets defined in the Profinet IO specification.

Errors occurring when configuring and setting the parameters of the fieldbus communicator and the connected pyrometer modules as well as external errors are reported by the communicator via channel specific diagnostic.

In productive data exchange between the I/O controller and the fieldbus communicator, one byte IOPS process data qualifiers are available for each module providing information of the validity of the pyrometer module data (good/ bad). In the event of an error occurs during operation, the problem-

indicator in APDU-Status is set by the communicator and a diagnostic alarm is additionally transmitted.

17.2.5.1 Error Bits of the Station Diagnosis (optional)

| Bit | Description |
|-----|--------------------------------------|
| 0 | Self test error |
| 1 | Box ambient temperature out of range |
| 2 | Sensing head communication error |
| 3 | Parameter error |
| 4 | Register write error |

17.2.5.2 Error Bits of the Pyrometer Module Diagnosis (optional)

| Bit | Description |
|-----|--|
| 0 | Object temperature out of range |
| 1 | Ambient (internal) temperature out of range |
| 2 | Parameter error |
| 3 | Register write error |
| 4 | Self test error |
| 5 | reserved |
| 6 | head registered but not connected → cable break? |

18 ASCII Programming

This section explains the system's ASCII communication protocols. A protocol is the set of commands that defines all possible communications with the sensor. The commands are described along with their associated ASCII command characters and related message format information. Use them when writing custom programs for your applications or when communicating with your sensor using a terminal program.

18.1 Transfer Modes

There are two possible transfer modes for the digital interface:

Poll Mode: By user-interface control, a parameter will be set or requested.

Burst Mode: A pre-defined data string ("burst string") will be transferred as fast as possible, as long as the burst mode is activated. The data will be transferred in one direction only, from the unit to the user interface.

V=P "P" starts the Poll mode (allows to request or to set parameters)

V=B "B" starts the Burst mode (data will be transferred as fast as possible; necessary: data string definition – "Burst string")

?X\$ gives the burst string parameters while in poll mode

Single Head System:

\$=UTIE "\$" sets the parameter combination ("burst string")

"U" unit (°C or °F)

"T" temperature value

"I" internal temperature of the sensing head

"E" emissivity

Multiple Head System:

\$=U1T1I1E2T2I2E "\$" sets the parameter combination ("burst string")

"U" unit (°C or °F)

"1T" temperature value for sensing head 1

"1I" internal temperature for sensing head 1

"1E" emissivity for sensing head 1

"2T" temperature value for sensing head 2

"2I" internal temperature for sensing head 2

"2E" emissivity for sensing head 2

Return from burst mode to poll mode:

If poll mode should activate while burst mode is still active, send a character and within the following few seconds the command V=P.

18.2 Command Structure

Requesting a parameter (Poll Mode)

?E<CR> "?" is the command for "Request"

"E" is the parameter requested

<CR> (carriage return, 0Dh) is closing the request.

Remark: It is possible to close with <CR> <LF>, 0Dh, 0Ah, but not necessary.

Setting a parameter (Poll Mode)

ASCII Programming

The parameter will be stored into the device EEPROM.

E=0.975<CR> "E" is the parameter to be set
"=" is the command for "set a parameter"
"0.975" is the value for the parameter
<CR> (carriage return, 0Dh) is closing the request
Remark: It is possible to close with <CR> <LF>, 0Dh, 0Ah, but not necessary.

Setting a parameter without writing into the EEPROM (Poll Mode)

This function is for test purposes only.

E#0.975<CR> "E" is the parameter to be set
"#" is the command for "set parameter without writing into the EEPROM"
"0.975" is the value for the parameter
<CR> (carriage return, 0Dh) is closing the request.
Remark: It is possible to close with <CR> <LF>, 0Dh, 0Ah, but not necessary.

Device response format:

!E0.975<CR><LF> "!" is the parameter for "Answer"
"E" is the parameter
"0.975" is the value for the parameter
<CR> <LF> (0Dh 0Ah) is closing the answer.

Error message

Syntax Error "" is the character for "Error"

18.3 Device Information

This information is factory installed, read only.

| Command | Description | Answer (Example) |
|---------|---------------------------------------|------------------|
| ?HI | Device name of the Head | "HIMI3LTS22" |
| ?HN | Serial Number of the Head | "HN98123" |
| ?XU | Device name of the Box | "XUMI3" |
| ?XV | Serial Number of the Box | "!XV0A0027" |
| ?XH | Maximum Temp. Range: e.g. for LT head | "!XH0600.0" |
| ?XB | Minimum Temp. Range: e.g. for LT head | "!XB-040.0" |

Table 15: Device Information

18.4 Device Setup

18.4.1 Temperature Calculation

U=C unit for the temperature value
E=0.950 Emissivity setting (Caution: according to the settings for "ES", see section 18.4.3 [Emissivity Setting and Alarm Set points](#), page 105.)
XG=1.000 Setting for transmission

For the calculation of the temperature value, it is possible to set an offset (relative number to be added to the temperature value), and a gain value.

DG=1.0000 Gain adjustment for the temperature signal
DO=0 Offset adjustment for the temperature signal

If the ambient temperature is not requested by the internal head temperature, you must set the ambient temperature values, as follows:

A=250.0 Ambient temperature (example)
AC=1 Control ambient background temp. compensation

18.4.2 Temperature Pre-Processing

The samples from the AD converter (energy values) can be processed before temperature calculation. The corresponding filter command is <FF>.

18.4.3 Emissivity Setting and Alarm Set points

The device allows three choices for the emissivity setting and two for the alarm output setting.

ES Selection of the emissivity setting.

ES=1 Emissivity set by a constant number according to the „E“ command
ES=E Emissivity set by a voltage on FTC1 (analog input)
ES=D Emissivity set by the entries in a table (selected by digital inputs FTC1 – FTC3)
?CE asks for the emissivity value that is actually used for temperature calculation

There are eight entries possible for emissivity setting ① and a related set point (threshold) ②. To be able to write or read these values, use the following commands:

EP=2 set pointer for table entry, e.g. to line 2 ③
EV=0.600 set the emissivity value for line 2 to 0.600 ④
SV=220.0 set the set point (threshold) for line 2 to 220.0 ⑤

| | Emissivity | Set Point |
|---|------------|-----------|
| 0 | 1,100 | 200,0 |
| 1 | 0,500 | 210,0 |
| 2 | 0,600 | 220,0 |
| 3 | 0,700 | 230,0 |
| 4 | 0,800 | 240,0 |
| 5 | 0,970 | 250,0 |
| 6 | 1,000 | 260,0 |
| 7 | 0,950 | 270,0 |

Figure 83: Table for Emissivity and Set Points

To activate these emissivity settings, you need to have the 3 external inputs (FTC) connected. According to the digital combination on the FTC wires, one of the table entries will be activated, see section 7.2 [Emissivity \(digital\)](#), page 41.

ASCII Programming

18.4.4 Post Processing

The following parameters can be set to determine the post processing mode. See section 8.5 [Post Processing](#), page 50.

P=5.0 peak hold, hold time: 5 s
F=12.5 valley hold, hold time: 12.5 s
G=10.0 averaging, average time (90%): 10 s
XY=3.0 advanced peak hold, hysteresis: 3 K
XY=-2.0 advanced valley hold, hysteresis: 2 K

Advanced Peak/Valley Hold with Averaging:

C=250.0 threshold: 250°C
AA=15.0 averaging time (90%): 15 s

18.5 Dynamic Data

All temperature related information is calculated multiple times per second. To request the dynamic data, the following commands are available:

?T target temperature
?I internal temperature of the sensing head
?XJ internal temperature of the electronics housing
?Q energy value of the infrared temperature
?XT trigger set point (active/inactive) for the FTC3 input

To check for resets (e.g. power shut down), use the command XI. Notice, after a reset, the unit is new initialized.

?XI asks for the reset status
!XI0 no reset occurred
!XI1 a reset occurred, new initialization of the unit
XI=0 sets the reset status back to 0

18.6 Device Control

18.6.1 Output for the Target Temperature

The signal output can be set to 4 – 20 mA, 0 – 20 mA or V. If current output is activated, the output can provide a predefined current:

XO2O=4 mode to 4 – 20 mA for output 2
O2O=13.57 output of a constant current at 13.57 mA at output 2
O2O=60 switches back to the temperature controlled output

18.6.2 Analog Output, Scaling

According to the temperature range of the model, it is possible to set a maximum voltage/current value according to a temperature value (e.g., the maximum current 20 mA shall represent 200°C / 392°F). The same setting is possible for the minimum value.

H2O=500 the maximum current/voltage value for output 2 is set to 500°C
L2O=0 the minimum current/voltage value for output 2 is set to 0°C

Remark: You cannot set this value for thermocouple output. The minimum span between the maximum / minimum settings is 20 K.

18.6.3 Alarm Output

The alarm output (see section 6.4 [Alarm Output RELAY](#), page 38) can be driven by two triggers:

- object temperature
- head ambient temperature


| | |
|----------|---|
| KH=off | no alarm control |
| KH=1 | object temperature drives alarm control |
| KH=2 | head ambient temperature drives alarm control |
| KB=0 | relay contacts permanently open |
| KB=1 | relay contacts permanently closed |
| KB=2 | relay contacts normally open |
| KB=3 | relay contacts normally closed |
| XS=125.3 | threshold setting to 125.3°C (if U=C is set) |



18.6.4 Factory default values

It is possible to reset the unit to the default values.

XF factory default values will be set

18.6.5 Lock Mode

The access to the unit is possible via serial interface (software) and via the direct user input (mode buttons, LCD display). It is possible to lock the  button. This allows the change of parameters to the unit only via software.

J=L direct user access to  button denied
J=U unlocking the  button

18.6.6 Mode Setting for the Digital Input FTC3

The digital input FTC3 (see section 7.4 [Trigger/Hold](#), page 43.) can be used as follows:

XN=T FTC3 as trigger
XN=H FTC3 with hold function

18.6.7 Ambient Background Temperature Compensation

To compensate the ambient background temperature, the following modes are available:

AC=0 no compensation
AC=1 compensation with a constant temperature value set with command A.
AC=2 compensation with an external voltage signal at the analog input FTC2 (0 V – 5V corresponds to low end and high end of temperature range). Current ambient temperature is readable with command A.

Note: The mode AC = 2 does not function in case of setting the command ES = D!

For more information regarding the ambient background temperature compensation feature, see section see section 7.3 [Ambient Temperature Compensation](#), page 41.

18.7 Addressing of Multiple Heads

The communication boxes support up to 8 heads. To direct a command to one head among the 8 possible, it is necessary to “address” the head command. Therefore, a number between 1 and 8 is set prior to the head command.

ASCII Programming

Examples for the head command "Emissivity":

?2E requests the emissivity for sensing head at address 2
 2E=0.975 sets the emissivity to 0.975 for sensing head at address 2
 !2E0.975 sensing head at address 2 confirms the emissivity setting

18.8 Addressing of Multiple Boxes (RS485)

Up to 32 communication boxes can be connected within an RS485 network. To direct a command to one box among the 32 possible, it is necessary to „address“ a command. Therefore, a 3-digit number is set prior the box command. The 3-digit number is determined between 001 and 032.

XA=024 will set address to 24 (box must not be in multidrop mode)

Changing an address:

(e.g. the address is changed from 17 to 24)

command answer
 „017?E“ „017E0.950“
 „017XA=024“ „017XA024“ setting of a new address
 „024?E“ „024E0.950“

Note: A box with the address 000 is a single box and not in multidrop mode.

If a command is transferred, starting with the 3-digit number 000, all boxes (with addresses from 001 to 032) connected will get this command – without to send an answer.

command answer
 “024?E” “024E0.950”
 “000E=0.5” will be executed from all units, no answer
 “024?E” “024E0.500”
 “012?E” “012E0.500”

18.9 Command Set

P ... Poll, B ... Burst, S ... Set, N ... Notification

n ... head number, v ... value, X ... uppercase letter

| Description | Char | Format | P | B | S | N | Legal Values | Factory default | Head / Box |
|---|------|--------------|---|---|---|---|--|-----------------|------------|
| Poll parameter | ? | ?X | √ | | | | ?T | | |
| Set parameter | = | X=... | | | √ | | BR=115200 | | |
| Error message | * | | | | | | *Syntax error | | |
| Acknowledge message | ! | | | | | | !BR115200 | | |
| Burst string format | \$ | ?\$ or \$=nT | √ | | √ | | | TIXJXT | |
| Ambient background temp. compensation | A | nA float | √ | √ | √ | | °C/°F | 23°C | H |
| Advanced hold with average | AA | nAA float | √ | √ | √ | | 0 – 999.9 s | 000.0s | H |
| Control ambient background temp. compensation | AC | nAC | √ | √ | √ | | 0 – sensor temp. 1 – via number 2 – via ext. input | 0 | H |

ASCII Programming

| Description | Char | Format | P | B | S | N | Legal Values | Factory default | Head / Box |
|--|------|--|---|---|---|---|--|-------------------|------------|
| Baud rate RS485 | BR | BR integer | √ | | √ | | 9600, 19200 38400, 57600, 115200 | 9600 | B |
| Advanced hold threshold | C | nC | √ | | √ | | | 300 °C/°F | H |
| Currently calculated emissivity | CE | nCE | √ | | | | 0.1 - 1.1 | | H |
| Cali Certification Date, Time | CFDT | nCFDT 32int 32int | √ | | | | CFDT=yyyymmdd hhmmss | | B |
| Cali Certification Low Temp | CFLT | nCFLT float float float float float | √ | | | | CFLT=23.0 SourceTemp1 MeasureTemp1 SourceTemp2 MeasureTemp2 SourceTemp3 MeasureTemp3 | | B |
| Cali Certification High Temp | CFHT | nCFHT float float float float | √ | | | | CFHT=80.0 SourceTemp1 MeasureTemp1 SourceTemp2 MeasureTemp2 | | B |
| Communication module | CM | ?CM | √ | | | | 0 – no module 1 – RS485 2 – Profibus 3 – Modbus 4 – reserved 5 – Ethernet 6 – Profinet | | B |
| Current calculation setpoint/ relay function | CS | nCS | √ | | | | °C/°F | | B |
| Sensor Gain | DG | nDG float | | | | | 0.8...1.2 | 1.0 | H |
| Sensor Offset | DO | nDO float | | | | | -200°C...+200°C | 0°C | H |
| Delete Sensor | DH | nDH | | | | | | | B |
| Box special | DS | string | √ | | √ | | e.g. !DSRAY (read only) | Set at production | B |
| Box Status Code | EC | EC 16 bit hex | √ | | | | hex value of Status Code ¹ | | B |
| External module | EM | ?EM | √ | | | | 0 – no module 2 – 2 channels 4 – 4 channels (analog outputs) | | B |

¹ Box Status Codes (read only):

| | |
|--------------------------------------|------|
| Self-test error | BIT0 |
| Box ambient temperature out of range | BIT1 |
| Sensing head communication error | BIT2 |
| Parameter error | BIT3 |
| Register write error | BIT4 |
| Analog module error | BIT5 |
| Profinet ready | BIT6 |

ASCII Programming

| Description | Char | Format | P | B | S | N | Legal Values | Factory default | Head / Box |
|-----------------------------------|------------|--------------------|---|---|-----|---|---|-----------------------------|------------|
| Presel. Emissivity Pointer | EP | nEP integer | | | | | 0 - 7 | 0 | H |
| Emissivity Source | ES | nES integer | √ | | √ | | I = Emissivity from Internal (by command) E = Emissivity from External analog input (0V - 5V) D = digital selected FTC1-3 | I | H |
| Presel. Emissivity Value | EV | nEV float | √ | | √ | | 0.1 - 1.1 | | H |
| Valley hold time ¹ | F | nF float | √ | √ | √ | √ | 0.0 - 998.9 s (999 = ∞) | 0.0 s | H |
| Flicker Filter | FF | nFF integer | | | | | 0...32768 | 0 (LT, G5) 3000 (1M, 2M) | H |
| Average time ² | G | nG float | √ | √ | √ | √ | 0 - 999.0 s | 0.0 s | H |
| Head Address | HA | nHA | √ | | | | 1 - 9 | | B |
| Connected heads | HC | string | √ | | | | e.g. !HC - no heads !HC1 2 3 7 8 | | B |
| Registered heads | HCR | string | √ | | √ | | e.g. !HCR - no heads !HCR1 2 3 7 8 HCR=0 -> new registration | | B |
| Head Status Code | HEC | nHEC 16 bit hex | √ | | | | hex value of Status Code ³ | | H |
| Head identification | HI | ?nHI string | √ | | (√) | | e.g. !7HIMIXLTS22 | set at production | H |
| Top temperature value of output 1 | H1O (H) | float | √ | | √ | √ | H1O= -40.0 or H= -40.0 | 500°C | B |
| Top temperature value of output 2 | H2O | float | √ | | √ | √ | H2O= -40.0 | 500°C | B |
| Laser Control ⁴ | HL | nHL integer | √ | | √ | √ | 0=off, 1=on, 2=flash, 3=external | 0 | H |
| Head serial number | HN | ?nHN integer | √ | | √ | | e.g. !nHN98123 | set at production | H |
| Head special | HS | ?nHS string | √ | | √ | | e.g. !99HSRAY | Set at production | H |
| Head Firmware Revision | HV | ?nHV | √ | | | | e.g. 1.01 | Set in FW | H |

¹ setting average / peak / valley / advanced hold cancels all other hold modes

² setting average / peak / valley / advanced hold cancels all other hold modes

³ Head Status Codes (read only):

| | |
|----------------------------------|--------------------------|
| Temperature unit | BIT0 // 0 = °C, 1 = °F |
| Object temperature out of range | BIT1 // 1 = out of range |
| Ambient temperature out of range | BIT2 // 1 = out of range |
| Parameter error for a command | BIT3 // 1 = error |
| Register write error | BIT4 // 1 = error |
| Self-test error | BIT5 // 1 = error |
| {reserved} | BIT6 |
| Ambient temperature compensation | BIT7 // 0 = off, 1 = on |

⁴ for MI3100 heads only

ASCII Programming

| Description | Char | Format | P | B | S | N | Legal Values | Factory default | Head / Box |
|---|------------|-------------------------------------|---|---|---|---|---|-----------------|------------|
| Restore Head Factory defaults | HXF | nHXF | | | √ | | | | H |
| Head ambient | I | ?nI float | √ | √ | | | answer !nI=99.9 | | H |
| Switch panel lock | J | X | √ | | √ | √ | L = locked U = unlocked | unlocked | B |
| Relay alarm output control depreciated for MI3 (use the commands KB and KH instead of) | K | K integer | √ | | √ | | 0 = off 1 = on 2 = Target norm. open 3 = Target norm. closed 4 = Intern norm. open 5 = Intern norm. closed | 2 | H |
| Relay alarm output control | KB | KB integer | √ | | √ | | 0 = off 1 = on 2 = norm. open 3 = norm. closed | 2 | B |
| Relay alarm output control | KH | nKH integer | √ | | √ | | 0 = off 1 = target temp. 2 = head ambient | 1 | H |
| Bottom temperature value of output 1 | L10 (L) | float | √ | | √ | √ | L10= -40.0 or L = -40.0 | 0°C | B |
| Bottom temperature value of output 2 | L20 | float | √ | | √ | √ | L20= -40.0 | 0°C | B |
| Output 1 source | O10 (O) | O10=1T float, or nT, or nI | √ | | √ | | v – float value n – head number if v = 60 – controlled by head 1(O10=1T) | O10=0 | B |
| Output 2 source | O20 | O20=1I float, or nT, or nI | √ | | √ | | v – float value n – head number if v = 60 – controlled by head 1(O20=1I) | O20=0 | B |
| Peak hold time ¹ | P | nP float | √ | √ | √ | √ | 0.0 - 998.9s (999 = ∞) | 0.0s | H |
| Power / AD value | Q | ?nQ | | | | | | | H |
| Presel. Setpoint | SV | nSV float | | | | | | 500°C | H |
| Target temperature | T | ?nT float | √ | √ | | | answer !nT=99.9 | | H |
| | | | | | | | | | |
| Get Analog Input 1 | TV1I | ?TV1I | | | | | 0 – 5V | | B |
| Get Analog Input 2 | TV2I | ?TV2I | | | | | 0 – 5V | | B |
| Temperature Unit | U | X | √ | √ | √ | √ | C / F | C | B/H |
| Poll / Burst mode | V | V=P | √ | | √ | | P=poll B=burst | poll mode | B |
| Burst string contents | X\$ | ?X\$ | √ | | | | | | B |
| Multidrop Address | XA | 0nn | √ | | √ | √ | 000 – 032 (000 --> single unit mode) | 000 | B |

¹ setting average / peak / valley / advanced hold cancels all other hold modes

ASCII Programming

| Description | Char | Format | P | B | S | N | Legal Values | Factory default | Head / Box |
|------------------------------|--------------|------------------|---|---|---|---|---|---|------------|
| Profibus / Modbus Address | XAS | n | √ | | √ | √ | 0 – 125 (Profibus) 1 – 247 (Modbus) | 0 (Profibus) 1 (Modbus) | B |
| Device Bottom range limit | XB | ?nXB float | √ | | √ | | answer !nXB=10.0 | | H |
| Restore Box Factory defaults | XF | | | | √ | | | except address | B |
| Transmission | XG | float | √ | √ | √ | | 0.1 – 1.0 | 1.0 | H |
| Device High range limit | XH | ?nXH float | √ | | √ | | read only | | H |
| Sensor initialization | XI | | | | | | 1 after reset 0 if XI=0 | 1 | B |
| Box Temperature | XJ | float | | | | | (°C/°F) | | B |
| FTC3 trigger/hold | XN | nXN | | | | | T = trigger, H = hold | T | H |
| analog Output 1 mode | XO10 (XO) | XO10=v (XO=v) | √ | | √ | | 0 – 0...20 mA ¹ 4 – 4...20 mA 5 – TCJ ² 6 – TCK 7 – TCR 8 – TCS 9 – 0...5 V 10 – 0...10 V 99 – disable (tristate) | XO10=9 for communication box (metal) XO10=99 for communication box 6TE, analog | B |
| analog Output 2 mode | XO20 | XO20=v | √ | | √ | | 0 – 0...20 mA 4 – 4...20 mA 9 – 0...5 V 10 – 0...10 V 99 – disable (tristate) | XO20=4 for communication box (metal) XO10=99 for communication box 6TE, analog | B |
| analog Output 3 mode | XO30 | XO30=v | √ | | √ | | 0 – 0...20 mA 4 – 4...20 mA 9 – 0...5 V 10 – 0...10 V 99 – disable (tristate) | XO30=99 for communication box 6TE, analog | |
| analog Output 4 mode | XO40 | XO40=v | √ | | √ | | 0 – 0...20 mA 4 – 4...20 mA 9 – 0...5 V 10 – 0...10 V 99 – disable (tristate) | XO40=99 for communication box 6TE, analog | |
| Box Firmware Revision | XR | ?XR | √ | | | | e.g. 1.01 | Set in FW | B |
| Setpoint relay function | XS | nXS float | √ | | √ | | °C/°F | 500°C | H |
| Trigger | XT | ?XT | √ | √ | | √ | 0 = inactive, 1 = active | 0 | B |
| Unit identification | XU | ?XU | √ | | | | e.g. !XUMI3COMM | set at production | |
| Serial number | XV | ?XV | √ | | | | e.g. !XV98123 | set at production | B |
| Advanced hold hysteresis | XY | nnnn | √ | | √ | | | | H |

¹ Mode 0 and 4 for MI3MCOMMA only

² Mode 5 to 8 for MI3COMM only

ASCII Programming

18.9.1 ASCII Commands for Ethernet and Profinet

| Description | Char | Format | P | B | S | N | Legal Values | Factory default | Head / Box | |
|--|------|--------|---|---|---|---|-------------------|--|------------|------|
| IP address | IP | | √ | | | | 192.xxx.xxx.xxx | 192.168.42.130 (Ethernet) 0.0.0.0 (Profinet) | B | 2.19 |
| Net Mask | NM | | √ | | | | 255.255.255.0 | 255.255.255.0 (Ethernet) 0.0.0.0 (Profinet) | B | 2.19 |
| Gateway | GW | | √ | | | | 192.xxx.xxx.xxx | 192.x.x.x (Ethernet) 0.0.0.0 (Profinet) | B | 2.19 |
| MAC address (read only) | MAC | | √ | | | | 001d8d 1xxxxx | 00:1d:8d:10:00:14 (Ethernet) 00:1d:8d:10:00:1f (Profinet) | B | 2.19 |
| PORT address (Ethernet only) | PORT | | √ | | | | 0 - 65535 | 6363 | B | 2.19 |
| DHCP on/off (Ethernet only) | IPU | | √ | | | | 0 - off 1 - on | 0 | B | 2.19 |
| Data logging on/off (Ethernet only) | DL | | √ | | | | 0 - off 1 - on | 0 | B | 2.19 |
| Data logging interval (Ethernet only) | DLI | | √ | | | | 1 - 2097120 s | 1s | B | 2.19 |
| Get Ethernet FW Version | ETV | | √ | | | | String, 4 char | 1.00 | B | 2.19 |
| Handle Profinet Name | PNN | | √ | | | | String, 10 char | commbx | B | 2.19 |
| Reset Ethernet module | RSE | | √ | | | | | | B | 2.19 |

Table 16: Command Set

19 Appendix

19.1 Determination of Emissivity

Emissivity is a measure of an object's ability to absorb and emit infrared energy. It can have a value between 0 and 1.0. For example a mirror has an emissivity of < 0.1 , while the so-called "Blackbody" reaches an emissivity value of 1.0. If a higher than actual emissivity value is set, the output will read low, provided the target temperature is above its ambient temperature. For example, if you have set 0.95 and the actual emissivity is 0.9, the temperature reading will be lower than the true temperature.

An object's emissivity can be determined by one of the following methods:

1. Determine the actual temperature of the material using an RTD (PT100), a thermocouple, or any other suitable contact temperature method. Next, measure the object's temperature and adjust emissivity setting until the correct temperature value is reached. This is the correct emissivity for the measured material.
2. For relatively low temperatures (up to 260°C / 500°F) place a plastic sticker (e.g. XXXRPMACED) on the object to be measured. This sticker should be large enough to cover the target spot. Next, measure the sticker's temperature using an emissivity setting of 0.95. Finally, measure the temperature of an adjacent area on the object and adjust the emissivity setting until the same temperature is reached. This is the correct emissivity for the measured material.
3. If possible, apply flat black paint to a portion of the surface of the object. The emissivity of the paint is 0.95. Next, measure the temperature of the painted area using an emissivity setting of 0.95. Finally, measure the temperature of an adjacent area on the object and adjust the emissivity until the same temperature is reached. This is the correct emissivity for the measured material.

19.2 Typical Emissivity Values

The following table provides a brief reference guide for determining emissivity and can be used when one of the above methods is not practical. Emissivity values shown in the table are only approximate, since several parameters may affect the emissivity of a material. These include the following:

- Temperature
- Angle of measurement
- Geometry (plane, concave, convex)
- Thickness
- Surface quality (polished, rough, oxidized, sandblasted)
- Spectral range of measurement
- Transmission (e.g. thin films plastics)

To optimize surface temperature measurements, consider the following guidelines:

- Determine the object's emissivity using the instrument which is also to be used for temperature measurements.
- Avoid reflections by shielding the object from surrounding temperature sources.
- For higher temperature objects, use instruments with the shortest wavelength possible.
- For translucent materials such as plastic foils or glass, assure that the background is uniform and lower in temperature than the object.
- Mount the instrument perpendicular to the surface, if possible. In all cases, do not exceed angles more than 30° from incidence.

Appendix

| Material | METALS | | |
|-----------------------|-------------------|-----------------|----------------------|
| | 3.9 μm | 5 μm | 8 – 14 μm |
| Aluminum | | | |
| Unoxidized | 0.02-0.2 | 0.02-0.2 | 0.02-0.1 |
| Oxidized | 0.2-0.4 | 0.2-0.4 | 0.2-0.4 |
| Alloy A3003, Oxidized | 0.4 | 0.4 | 0.3 |
| Roughened | 0.1-0.4 | 0.1-0.4 | 0.1-0.3 |
| Polished | 0.02-0.1 | 0.02-0.1 | 0.02-0.1 |
| Brass | | | |
| Polished | 0.01-0.05 | 0.01-0.05 | 0.01-0.05 |
| Burnished | 0.3 | 0.3 | 0.3 |
| Oxidized | 0.5 | 0.5 | 0.5 |
| Chromium | 0.03-0.3 | 0.03-0.3 | 0.02-0.2 |
| Copper | | | |
| Polished | 0.03 | 0.03 | 0.03 |
| Roughened | 0.05-0.15 | 0.05-0.15 | 0.05-0.1 |
| Oxidized | 0.5-0.8 | 0.5-0.8 | 0.4-0.8 |
| Gold | 0.01-0.1 | 0.01-0.1 | 0.01-0.1 |
| Haynes | | | |
| Alloy | 0.3-0.8 | 0.3-0.8 | 0.3-0.8 |
| Inconel | | | |
| Oxidized | 0.6-0.9 | 0.6-0.9 | 0.7-0.95 |
| Sandblasted | 0.3-0.6 | 0.3-0.6 | 0.3-0.6 |
| Electropolished | 0.15 | 0.15 | 0.15 |
| Iron | | | |
| Oxidized | 0.6-0.9 | 0.6-0.9 | 0.5-0.9 |
| Unoxidized | 0.05-0.25 | 0.05-0.25 | 0.05-0.2 |
| Rusted | 0.5-0.8 | 0.5-0.8 | 0.5-0.7 |
| Molten | — | — | — |
| Iron, Cast | | | |
| Oxidized | 0.65-0.95 | 0.65-0.95 | 0.6-0.95 |
| Unoxidized | 0.25 | 0.25 | 0.2 |
| Molten | 0.2-0.3 | 0.2-0.3 | 0.2-0.3 |
| Iron, Wrought | | | |
| Dull | 0.9 | 0.9 | 0.9 |
| Lead | | | |
| Polished | 0.05-0.2 | 0.05-0.2 | 0.05-0.1 |
| Rough | 0.4 | 0.4 | 0.4 |
| Oxidized | 0.2-0.7 | 0.2-0.7 | 0.2-0.6 |
| Magnesium | 0.03-0.15 | 0.03-0.15 | 0.02-0.1 |
| Mercury | 0.05-0.15 | 0.05-0.15 | 0.05-0.15 |
| Molybdenum | | | |
| Oxidized | 0.3-0.7 | 0.3-0.7 | 0.2-0.6 |
| Unoxidized | 0.1-0.15 | 0.1-0.15 | 0.1 |
| Monel (Ni-Cu) | 0.1-0.5 | 0.1-0.5 | 0.1-0.14 |
| Nickel | | | |
| Oxidized | 0.3-0.6 | 0.3-0.6 | 0.2-0.5 |
| Electrolytic | 0.1-0.15 | 0.1-0.15 | 0.05-0.15 |
| Platinum | | | |
| Black | 0.9 | 0.9 | 0.9 |
| Silver | 0.02 | 0.02 | 0.02 |
| Steel | | | |
| Cold-Rolled | 0.8-0.9 | 0.8-0.9 | 0.7-0.9 |
| Ground Sheet | 0.5-0.7 | 0.5-0.7 | 0.4-0.6 |

| | | | |
|------------------|-----------|-----------|----------|
| Polished Sheet | 0.1 | 0.1 | 0.1 |
| Molten | 0.1-0.2 | 0.1-0.2 | — |
| Oxidized | 0.7-0.9 | 0.7-0.9 | 0.7-0.9 |
| Stainless | 0.15-0.8 | 0.15-0.8 | 0.1-0.8 |
| Tin (Unoxidized) | 0.05 | 0.05 | 0.05 |
| Titanium | | | |
| Polished | 0.1-0.3 | 0.1-0.3 | 0.05-0.2 |
| Oxidized | 0.5-0.7 | 0.5-0.7 | 0.5-0.6 |
| Tungsten | 0.05-0.5 | 0.05-0.5 | 0.03 |
| Polished | 0.05-0.25 | 0.05-0.25 | 0.03-0.1 |
| Zinc | | | |
| Oxidized | 0.1 | 0.1 | 0.1 |
| Polished | 0.03 | 0.03 | 0.02 |

Tab. 17: Typical Emissivity Values for Metals

Appendix

| Material | NON-METALS | | |
|--|-------------------|-----------------|----------------------|
| | 3.9 μm | 5 μm | 8 – 14 μm |
| Asbestos | | 0.9 | 0.95 |
| Asphalt | | 0.95 | 0.95 |
| Basalt | | 0.7 | 0.7 |
| Carbon | | | |
| Unoxidized | | 0.8-0.9 | 0.8-0.9 |
| Graphite | | 0.7-0.9 | 0.7-0.8 |
| Carborundum | | 0.9 | 0.9 |
| Ceramic | | 0.8-0.95 | 0.95 |
| Clay | | 0.85-0.95 | 0.95 |
| Concrete | | 0.9 | 0.95 |
| Cloth | | 0.95 | 0.95 |
| Glass | | | |
| Plate | | 0.98 | 0.85 |
| " Gob" | | 0.9 | — |
| Gravel | | 0.95 | 0.95 |
| Gypsum | | 0.4-0.97 | 0.8-0.95 |
| Ice | | — | 0.98 |
| Limestone | | 0.4-0.98 | 0.98 |
| Paint (non-al.) | | — | 0.9-0.95 |
| Paper (any color) | | 0.95 | 0.95 |
| Plastic, greater than 500 μm (0.02 in) thickness | | 0.95 | 0.95 |
| Rubber | | 0.9 | 0.95 |
| Sand | | 0.9 | 0.9 |
| Snow | | — | 0.9 |
| Soil | | — | 0.9-0.98 |
| Water | | — | 0.93 |
| Wood, Natural | | 0.9-0.95 | 0.9-0.95 |

Tab. 18: Typical Emissivity Values for Non-Metals

20 Notes